

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Continuing Calib Date: 10/17/2024 Initial Calibration Date(s): 10/08/2024 10/08/2024

Continuing Calib Time: 02:23 Initial Calibration Time(s): 16:30 23:46

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.92	5.92	5.82	6.02	0.00
Aroclor-1016-2	(2)	5.94	5.94	5.84	6.04	0.00
Aroclor-1016-3	(3)	6.00	6.01	5.91	6.11	0.01
Aroclor-1016-4	(4)	6.10	6.10	6.00	6.20	0.00
Aroclor-1016-5	(5)	6.40	6.40	6.30	6.50	0.00
Aroclor-1260-1	(1)	7.52	7.52	7.42	7.62	0.00
Aroclor-1260-2	(2)	7.77	7.78	7.68	7.88	0.01
Aroclor-1260-3	(3)	8.13	8.14	8.04	8.24	0.01
Aroclor-1260-4	(4)	8.37	8.38	8.28	8.48	0.01
Aroclor-1260-5	(5)	8.71	8.71	8.61	8.81	0.00
Tetrachloro-m-xylene		4.75	4.76	4.66	4.86	0.01
Decachlorobiphenyl		10.66	10.67	10.57	10.77	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Continuing Calib Date: 10/17/2024 Initial Calibration Date(s): 10/08/2024 10/08/2024

Continuing Calib Time: 02:23 Initial Calibration Time(s): 16:30 23:46

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.16	5.16	5.06	5.26	0.00
Aroclor-1016-2	(2)	5.18	5.18	5.08	5.28	0.00
Aroclor-1016-3	(3)	5.36	5.36	5.26	5.46	0.00
Aroclor-1016-4	(4)	5.40	5.40	5.30	5.50	0.00
Aroclor-1016-5	(5)	5.62	5.62	5.52	5.72	0.00
Aroclor-1260-1	(1)	6.66	6.66	6.56	6.76	0.00
Aroclor-1260-2	(2)	6.85	6.85	6.75	6.95	0.00
Aroclor-1260-3	(3)	7.00	7.01	6.91	7.11	0.01
Aroclor-1260-4	(4)	7.48	7.48	7.38	7.58	0.00
Aroclor-1260-5	(5)	7.72	7.72	7.62	7.82	0.00
Tetrachloro-m-xylene		4.05	4.05	3.95	4.15	0.00
Decachlorobiphenyl		9.21	9.22	9.12	9.32	0.01