

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Continuing Calib Date: 10/24/2024 Initial Calibration Date(s): 10/21/2024 10/21/2024

Continuing Calib Time: 14:13 Initial Calibration Time(s): 12:35 15:15

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|------|------------|
| Aroclor-1016-1       | (1) | 4.42       | 4.42      | 4.32                   | 4.52 | 0.00       |
| Aroclor-1016-2       | (2) | 4.44       | 4.44      | 4.34                   | 4.54 | 0.00       |
| Aroclor-1016-3       | (3) | 4.50       | 4.49      | 4.39                   | 4.59 | 0.00       |
| Aroclor-1016-4       | (4) | 4.59       | 4.58      | 4.48                   | 4.68 | -0.01      |
| Aroclor-1016-5       | (5) | 4.87       | 4.87      | 4.77                   | 4.97 | 0.00       |
| Aroclor-1260-1       | (1) | 5.97       | 5.96      | 5.86                   | 6.06 | -0.01      |
| Aroclor-1260-2       | (2) | 6.23       | 6.22      | 6.12                   | 6.32 | 0.00       |
| Aroclor-1260-3       | (3) | 6.58       | 6.57      | 6.47                   | 6.67 | -0.01      |
| Aroclor-1260-4       | (4) | 6.79       | 6.79      | 6.69                   | 6.89 | 0.00       |
| Aroclor-1260-5       | (5) | 7.11       | 7.10      | 7.00                   | 7.20 | -0.01      |
| Tetrachloro-m-xylene |     | 3.33       | 3.33      | 3.23                   | 3.43 | 0.00       |
| Decachlorobiphenyl   |     | 8.46       | 8.45      | 8.35                   | 8.55 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Continuing Calib Date: 10/24/2024 Initial Calibration Date(s): 10/21/2024 10/21/2024

Continuing Calib Time: 14:13 Initial Calibration Time(s): 12:35 15:15

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 3.73       | 3.73      | 3.63      | 3.83 | 0.00       |
| Aroclor-1016-2       | (2) | 3.75       | 3.75      | 3.65      | 3.85 | 0.01       |
| Aroclor-1016-3       | (3) | 3.90       | 3.90      | 3.80      | 4.00 | 0.00       |
| Aroclor-1016-4       | (4) | 3.95       | 3.95      | 3.85      | 4.05 | 0.00       |
| Aroclor-1016-5       | (5) | 4.15       | 4.15      | 4.05      | 4.25 | 0.00       |
| Aroclor-1260-1       | (1) | 5.13       | 5.14      | 5.04      | 5.24 | 0.01       |
| Aroclor-1260-2       | (2) | 5.32       | 5.32      | 5.22      | 5.42 | 0.00       |
| Aroclor-1260-3       | (3) | 5.46       | 5.47      | 5.37      | 5.57 | 0.01       |
| Aroclor-1260-4       | (4) | 5.92       | 5.92      | 5.82      | 6.02 | 0.00       |
| Aroclor-1260-5       | (5) | 6.17       | 6.17      | 6.07      | 6.27 | 0.00       |
| Tetrachloro-m-xylene |     | 2.74       | 2.74      | 2.64      | 2.84 | 0.00       |
| Decachlorobiphenyl   |     | 7.48       | 7.48      | 7.38      | 7.58 | 0.00       |