

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Continuing Calib Date: 10/24/2024 Initial Calibration Date(s): 10/21/2024 10/21/2024

Continuing Calib Time: 19:21 Initial Calibration Time(s): 12:35 15:15

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.42	4.42	4.32	4.52	0.00
Aroclor-1016-2	(2)	4.43	4.44	4.34	4.54	0.01
Aroclor-1016-3	(3)	4.49	4.49	4.39	4.59	0.00
Aroclor-1016-4	(4)	4.58	4.58	4.48	4.68	0.00
Aroclor-1016-5	(5)	4.87	4.87	4.77	4.97	0.00
Aroclor-1260-1	(1)	5.96	5.96	5.86	6.06	0.00
Aroclor-1260-2	(2)	6.22	6.22	6.12	6.32	0.00
Aroclor-1260-3	(3)	6.57	6.57	6.47	6.67	0.00
Aroclor-1260-4	(4)	6.79	6.79	6.69	6.89	0.00
Aroclor-1260-5	(5)	7.10	7.10	7.00	7.20	0.00
Tetrachloro-m-xylene		3.33	3.33	3.23	3.43	0.00
Decachlorobiphenyl		8.45	8.45	8.35	8.55	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Continuing Calib Date: 10/24/2024 Initial Calibration Date(s): 10/21/2024 10/21/2024

Continuing Calib Time: 19:21 Initial Calibration Time(s): 12:35 15:15

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	3.73	3.73	3.63	3.83	0.00
Aroclor-1016-2	(2)	3.75	3.75	3.65	3.85	0.01
Aroclor-1016-3	(3)	3.90	3.90	3.80	4.00	0.00
Aroclor-1016-4	(4)	3.95	3.95	3.85	4.05	0.00
Aroclor-1016-5	(5)	4.15	4.15	4.05	4.25	0.00
Aroclor-1260-1	(1)	5.13	5.14	5.04	5.24	0.01
Aroclor-1260-2	(2)	5.32	5.32	5.22	5.42	0.00
Aroclor-1260-3	(3)	5.46	5.47	5.37	5.57	0.01
Aroclor-1260-4	(4)	5.92	5.92	5.82	6.02	0.00
Aroclor-1260-5	(5)	6.17	6.17	6.07	6.27	0.00
Tetrachloro-m-xylene		2.74	2.74	2.64	2.84	0.00
Decachlorobiphenyl		7.47	7.48	7.38	7.58	0.01