

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Continuing Calib Date: 10/16/2024 Initial Calibration Date(s): 10/16/2024 10/16/2024

Continuing Calib Time: 14:31 Initial Calibration Time(s): 07:37 13:14

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.88	4.87	4.77	4.97	-0.01
Aroclor-1016-2	(2)	4.90	4.89	4.79	4.99	-0.01
Aroclor-1016-3	(3)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-4	(4)	5.07	5.06	4.96	5.16	-0.01
Aroclor-1016-5	(5)	5.38	5.37	5.27	5.47	-0.01
Aroclor-1260-1	(1)	6.59	6.58	6.48	6.68	-0.01
Aroclor-1260-2	(2)	6.88	6.87	6.77	6.97	-0.01
Aroclor-1260-3	(3)	7.27	7.26	7.16	7.36	-0.01
Aroclor-1260-4	(4)	7.51	7.50	7.40	7.60	-0.01
Aroclor-1260-5	(5)	7.85	7.84	7.74	7.94	-0.01
Tetrachloro-m-xylene		3.65	3.65	3.55	3.75	0.00
Decachlorobiphenyl		9.55	9.54	9.44	9.64	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Continuing Calib Date: 10/16/2024 Initial Calibration Date(s): 10/16/2024 10/16/2024

Continuing Calib Time: 14:31 Initial Calibration Time(s): 07:37 13:14

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.10	4.10	4.00	4.20	0.00
Aroclor-1016-2	(2)	4.12	4.12	4.02	4.22	0.00
Aroclor-1016-3	(3)	4.30	4.30	4.20	4.40	0.00
Aroclor-1016-4	(4)	4.35	4.35	4.25	4.45	0.00
Aroclor-1016-5	(5)	4.57	4.57	4.47	4.67	0.00
Aroclor-1260-1	(1)	5.67	5.67	5.57	5.77	0.00
Aroclor-1260-2	(2)	5.88	5.88	5.78	5.98	0.00
Aroclor-1260-3	(3)	6.04	6.04	5.94	6.14	0.00
Aroclor-1260-4	(4)	6.55	6.55	6.45	6.65	0.00
Aroclor-1260-5	(5)	6.81	6.81	6.71	6.91	0.00
Tetrachloro-m-xylene		2.98	2.98	2.88	3.08	0.00
Decachlorobiphenyl		8.27	8.27	8.17	8.37	0.00