

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL01 Date Analyzed: 10/14/2024

Lab Sample No.: AR1660CCC500 Data File : PO107094.D Time Analyzed: 09:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.523	5.424	5.624	467.260	500.000	-6.5
Aroclor-1016-2	5.545	5.446	5.646	476.450	500.000	-4.7
Aroclor-1016-3	5.608	5.509	5.709	496.410	500.000	-0.7
Aroclor-1016-4	5.704	5.605	5.805	494.270	500.000	-1.1
Aroclor-1016-5	5.999	5.899	6.099	499.530	500.000	-0.1
Aroclor-1260-1	7.126	7.024	7.224	513.000	500.000	2.6
Aroclor-1260-2	7.382	7.281	7.481	500.370	500.000	0.1
Aroclor-1260-3	7.743	7.641	7.841	476.660	500.000	-4.7
Aroclor-1260-4	7.968	7.865	8.065	466.840	500.000	-6.6
Aroclor-1260-5	8.281	8.179	8.379	471.000	500.000	-5.8
Decachlorobiphenyl	10.059	9.956	10.156	55.990	50.000	12.0
Tetrachloro-m-xylene	4.374	4.276	4.476	43.740	50.000	-12.5

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL01 Date Analyzed: 10/14/2024

Lab Sample No.: AR1660CCC500 Data File : PO107094.D Time Analyzed: 09:38

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.728	4.631	4.831	460.120	500.000	-8.0
Aroclor-1016-2	4.747	4.650	4.850	458.170	500.000	-8.4
Aroclor-1016-3	4.923	4.826	5.026	475.690	500.000	-4.9
Aroclor-1016-4	4.964	4.867	5.067	479.670	500.000	-4.1
Aroclor-1016-5	5.177	5.081	5.281	477.620	500.000	-4.5
Aroclor-1260-1	6.210	6.113	6.313	489.160	500.000	-2.2
Aroclor-1260-2	6.396	6.300	6.500	487.030	500.000	-2.6
Aroclor-1260-3	6.550	6.453	6.653	494.820	500.000	-1.0
Aroclor-1260-4	7.020	6.924	7.124	498.600	500.000	-0.3
Aroclor-1260-5	7.261	7.164	7.364	503.180	500.000	0.6
Decachlorobiphenyl	8.643	8.547	8.747	48.670	50.000	-2.7
Tetrachloro-m-xylene	3.646	3.548	3.748	44.560	50.000	-10.9