

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL02 Date Analyzed: 10/14/2024

Lab Sample No.: AR1660CCC500 Data File : PO107105.D Time Analyzed: 13:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.523	5.424	5.624	461.190	500.000	-7.8
Aroclor-1016-2	5.546	5.446	5.646	468.330	500.000	-6.3
Aroclor-1016-3	5.608	5.509	5.709	483.890	500.000	-3.2
Aroclor-1016-4	5.705	5.605	5.805	480.180	500.000	-4.0
Aroclor-1016-5	6.000	5.899	6.099	486.530	500.000	-2.7
Aroclor-1260-1	7.126	7.024	7.224	498.420	500.000	-0.3
Aroclor-1260-2	7.383	7.281	7.481	487.000	500.000	-2.6
Aroclor-1260-3	7.744	7.641	7.841	463.660	500.000	-7.3
Aroclor-1260-4	7.969	7.865	8.065	439.840	500.000	-12.0
Aroclor-1260-5	8.282	8.179	8.379	452.300	500.000	-9.5
Decachlorobiphenyl	10.063	9.956	10.156	54.740	50.000	9.5
Tetrachloro-m-xylene	4.375	4.276	4.476	43.520	50.000	-13.0

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL02 Date Analyzed: 10/14/2024

Lab Sample No.: AR1660CCC500 Data File : PO107105.D Time Analyzed: 13:33

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.727	4.631	4.831	469.510	500.000	-6.1
Aroclor-1016-2	4.746	4.650	4.850	472.690	500.000	-5.5
Aroclor-1016-3	4.923	4.826	5.026	496.210	500.000	-0.8
Aroclor-1016-4	4.964	4.867	5.067	488.420	500.000	-2.3
Aroclor-1016-5	5.177	5.081	5.281	483.750	500.000	-3.3
Aroclor-1260-1	6.208	6.113	6.313	497.800	500.000	-0.4
Aroclor-1260-2	6.396	6.300	6.500	499.040	500.000	-0.2
Aroclor-1260-3	6.550	6.453	6.653	499.380	500.000	-0.1
Aroclor-1260-4	7.021	6.924	7.124	504.930	500.000	1.0
Aroclor-1260-5	7.260	7.164	7.364	515.210	500.000	3.0
Decachlorobiphenyl	8.641	8.547	8.747	49.870	50.000	-0.3
Tetrachloro-m-xylene	3.646	3.548	3.748	43.590	50.000	-12.8