

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL01 Date Analyzed: 10/16/2024

Lab Sample No.: AR1660CCC500 Data File : PP067889.D Time Analyzed: 09:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.921	5.819	6.019	516.540	500.000	3.3
Aroclor-1016-2	5.943	5.842	6.042	497.770	500.000	-0.4
Aroclor-1016-3	6.006	5.905	6.105	520.300	500.000	4.1
Aroclor-1016-4	6.104	6.003	6.203	491.490	500.000	-1.7
Aroclor-1016-5	6.399	6.298	6.498	493.390	500.000	-1.3
Aroclor-1260-1	7.522	7.422	7.622	461.920	500.000	-7.6
Aroclor-1260-2	7.775	7.675	7.875	459.990	500.000	-8.0
Aroclor-1260-3	8.137	8.037	8.237	450.140	500.000	-10.0
Aroclor-1260-4	8.376	8.275	8.475	464.780	500.000	-7.0
Aroclor-1260-5	8.713	8.612	8.812	448.980	500.000	-10.2
Decachlorobiphenyl	10.672	10.569	10.769	47.670	50.000	-4.7
Tetrachloro-m-xylene	4.757	4.655	4.855	50.720	50.000	1.4

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL01 Date Analyzed: 10/16/2024

Lab Sample No.: AR1660CCC500 Data File : PP067889.D Time Analyzed: 09:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.160	5.061	5.261	478.820	500.000	-4.2
Aroclor-1016-2	5.180	5.081	5.281	475.150	500.000	-5.0
Aroclor-1016-3	5.360	5.262	5.462	485.720	500.000	-2.9
Aroclor-1016-4	5.400	5.301	5.501	479.970	500.000	-4.0
Aroclor-1016-5	5.619	5.520	5.720	473.720	500.000	-5.3
Aroclor-1260-1	6.662	6.564	6.764	453.610	500.000	-9.3
Aroclor-1260-2	6.848	6.750	6.950	457.410	500.000	-8.5
Aroclor-1260-3	7.006	6.908	7.108	458.680	500.000	-8.3
Aroclor-1260-4	7.481	7.383	7.583	450.990	500.000	-9.8
Aroclor-1260-5	7.719	7.622	7.822	453.080	500.000	-9.4
Decachlorobiphenyl	9.215	9.119	9.319	44.890	50.000	-10.2
Tetrachloro-m-xylene	4.052	3.952	4.152	47.890	50.000	-4.2