

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL02 Date Analyzed: 10/16/2024

Lab Sample No.: AR1660CCC500 Data File : PP067900.D Time Analyzed: 12:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.918	5.819	6.019	507.820	500.000	1.6
Aroclor-1016-2	5.941	5.842	6.042	499.470	500.000	-0.1
Aroclor-1016-3	6.004	5.905	6.105	499.090	500.000	-0.2
Aroclor-1016-4	6.102	6.003	6.203	484.790	500.000	-3.0
Aroclor-1016-5	6.396	6.298	6.498	499.270	500.000	-0.1
Aroclor-1260-1	7.520	7.422	7.622	472.880	500.000	-5.4
Aroclor-1260-2	7.773	7.675	7.875	470.400	500.000	-5.9
Aroclor-1260-3	8.135	8.037	8.237	471.700	500.000	-5.7
Aroclor-1260-4	8.373	8.275	8.475	463.510	500.000	-7.3
Aroclor-1260-5	8.710	8.612	8.812	469.580	500.000	-6.1
Decachlorobiphenyl	10.667	10.569	10.769	48.370	50.000	-3.3
Tetrachloro-m-xylene	4.754	4.655	4.855	50.030	50.000	0.1

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL02 Date Analyzed: 10/16/2024

Lab Sample No.: AR1660CCC500 Data File : PP067900.D Time Analyzed: 12:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.158	5.061	5.261	466.270	500.000	-6.7
Aroclor-1016-2	5.178	5.081	5.281	468.510	500.000	-6.3
Aroclor-1016-3	5.359	5.262	5.462	473.320	500.000	-5.3
Aroclor-1016-4	5.399	5.301	5.501	474.180	500.000	-5.2
Aroclor-1016-5	5.617	5.520	5.720	468.920	500.000	-6.2
Aroclor-1260-1	6.661	6.564	6.764	459.910	500.000	-8.0
Aroclor-1260-2	6.847	6.750	6.950	464.710	500.000	-7.1
Aroclor-1260-3	7.005	6.908	7.108	461.920	500.000	-7.6
Aroclor-1260-4	7.479	7.383	7.583	459.480	500.000	-8.1
Aroclor-1260-5	7.718	7.622	7.822	458.120	500.000	-8.4
Decachlorobiphenyl	9.212	9.119	9.319	45.690	50.000	-8.6
Tetrachloro-m-xylene	4.050	3.952	4.152	47.290	50.000	-5.4