

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/21/2024 10/21/2024

Client Sample No.: CCAL02 Date Analyzed: 10/24/2024

Lab Sample No.: AR1660CCC500 Data File : PQ069337.D Time Analyzed: 19:21

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.416	4.315	4.515	516.970	500.000	3.4
Aroclor-1016-2	4.434	4.335	4.535	450.390	500.000	-9.9
Aroclor-1016-3	4.491	4.392	4.592	450.110	500.000	-10.0
Aroclor-1016-4	4.582	4.483	4.683	444.710	500.000	-11.1
Aroclor-1016-5	4.866	4.767	4.967	454.710	500.000	-9.1
Aroclor-1260-1	5.961	5.862	6.062	451.300	500.000	-9.7
Aroclor-1260-2	6.221	6.121	6.321	442.270	500.000	-11.5
Aroclor-1260-3	6.573	6.474	6.674	424.760	500.000	-15.0
Aroclor-1260-4	6.789	6.690	6.890	410.010	500.000	-18.0
Aroclor-1260-5	7.100	7.000	7.200	413.000	500.000	-17.4
Decachlorobiphenyl	8.453	8.352	8.552	40.680	50.000	-18.6
Tetrachloro-m-xylene	3.328	3.228	3.428	44.270	50.000	-11.5

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/21/2024 10/21/2024

Client Sample No.: CCAL02 Date Analyzed: 10/24/2024

Lab Sample No.: AR1660CCC500 Data File : PQ069337.D Time Analyzed: 19:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.730	3.630	3.830	532.130	500.000	6.4
Aroclor-1016-2	3.745	3.646	3.846	534.400	500.000	6.9
Aroclor-1016-3	3.904	3.804	4.004	527.370	500.000	5.5
Aroclor-1016-4	3.953	3.853	4.053	546.900	500.000	9.4
Aroclor-1016-5	4.146	4.047	4.247	510.230	500.000	2.0
Aroclor-1260-1	5.134	5.035	5.235	500.240	500.000	0.0
Aroclor-1260-2	5.323	5.224	5.424	540.270	500.000	8.1
Aroclor-1260-3	5.463	5.365	5.565	510.940	500.000	2.2
Aroclor-1260-4	5.922	5.823	6.023	501.670	500.000	0.3
Aroclor-1260-5	6.168	6.069	6.269	502.910	500.000	0.6
Decachlorobiphenyl	7.474	7.377	7.577	53.380	50.000	6.8
Tetrachloro-m-xylene	2.739	2.639	2.839	50.470	50.000	0.9