

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/16/2024 10/16/2024

Client Sample No.: CCAL01 Date Analyzed: 10/16/2024

Lab Sample No.: AR1660CCC500 Data File : PR068715.D Time Analyzed: 14:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.878	4.769	4.969	484.590	500.000	-3.1
Aroclor-1016-2	4.900	4.791	4.991	481.190	500.000	-3.8
Aroclor-1016-3	4.964	4.855	5.055	492.890	500.000	-1.4
Aroclor-1016-4	5.066	4.957	5.157	488.010	500.000	-2.4
Aroclor-1016-5	5.381	5.272	5.472	494.300	500.000	-1.1
Aroclor-1260-1	6.593	6.484	6.684	492.010	500.000	-1.6
Aroclor-1260-2	6.876	6.766	6.966	477.990	500.000	-4.4
Aroclor-1260-3	7.266	7.157	7.357	497.610	500.000	-0.5
Aroclor-1260-4	7.509	7.399	7.599	484.120	500.000	-3.2
Aroclor-1260-5	7.846	7.737	7.937	471.210	500.000	-5.8
Decachlorobiphenyl	9.553	9.441	9.641	47.970	50.000	-4.1
Tetrachloro-m-xylene	3.654	3.547	3.747	49.150	50.000	-1.7

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/16/2024 10/16/2024

Client Sample No.: CCAL01 Date Analyzed: 10/16/2024

Lab Sample No.: AR1660CCC500 Data File : PR068715.D Time Analyzed: 14:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.104	4.003	4.203	480.320	500.000	-3.9
Aroclor-1016-2	4.122	4.022	4.222	495.000	500.000	-1.0
Aroclor-1016-3	4.302	4.202	4.402	493.250	500.000	-1.4
Aroclor-1016-4	4.353	4.253	4.453	501.630	500.000	0.3
Aroclor-1016-5	4.572	4.472	4.672	497.710	500.000	-0.5
Aroclor-1260-1	5.672	5.572	5.772	489.260	500.000	-2.1
Aroclor-1260-2	5.878	5.778	5.978	482.600	500.000	-3.5
Aroclor-1260-3	6.038	5.937	6.137	490.330	500.000	-1.9
Aroclor-1260-4	6.547	6.447	6.647	485.170	500.000	-3.0
Aroclor-1260-5	6.813	6.711	6.911	480.580	500.000	-3.9
Decachlorobiphenyl	8.274	8.171	8.371	48.010	50.000	-4.0
Tetrachloro-m-xylene	2.978	2.878	3.078	48.790	50.000	-2.4