

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/16/2024 10/16/2024

Client Sample No.: CCAL02 Date Analyzed: 10/16/2024

Lab Sample No.: AR1660CCC500 Data File : PR068726.D Time Analyzed: 17:12

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.869	4.769	4.969	505.510	500.000	1.1
Aroclor-1016-2	4.891	4.791	4.991	489.290	500.000	-2.1
Aroclor-1016-3	4.955	4.855	5.055	501.210	500.000	0.2
Aroclor-1016-4	5.057	4.957	5.157	498.710	500.000	-0.3
Aroclor-1016-5	5.372	5.272	5.472	500.600	500.000	0.1
Aroclor-1260-1	6.584	6.484	6.684	498.970	500.000	-0.2
Aroclor-1260-2	6.866	6.766	6.966	496.180	500.000	-0.8
Aroclor-1260-3	7.257	7.157	7.357	505.570	500.000	1.1
Aroclor-1260-4	7.499	7.399	7.599	498.320	500.000	-0.3
Aroclor-1260-5	7.837	7.737	7.937	497.960	500.000	-0.4
Decachlorobiphenyl	9.539	9.441	9.641	52.530	50.000	5.1
Tetrachloro-m-xylene	3.647	3.547	3.747	50.140	50.000	0.3

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/16/2024 10/16/2024

Client Sample No.: CCAL02 Date Analyzed: 10/16/2024

Lab Sample No.: AR1660CCC500 Data File : PR068726.D Time Analyzed: 17:12

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.103	4.003	4.203	501.350	500.000	0.3
Aroclor-1016-2	4.122	4.022	4.222	509.890	500.000	2.0
Aroclor-1016-3	4.302	4.202	4.402	512.090	500.000	2.4
Aroclor-1016-4	4.353	4.253	4.453	517.420	500.000	3.5
Aroclor-1016-5	4.572	4.472	4.672	514.480	500.000	2.9
Aroclor-1260-1	5.670	5.572	5.772	503.530	500.000	0.7
Aroclor-1260-2	5.877	5.778	5.978	500.480	500.000	0.1
Aroclor-1260-3	6.036	5.937	6.137	507.880	500.000	1.6
Aroclor-1260-4	6.544	6.447	6.647	503.400	500.000	0.7
Aroclor-1260-5	6.810	6.711	6.911	501.580	500.000	0.3
Decachlorobiphenyl	8.269	8.171	8.371	50.880	50.000	1.8
Tetrachloro-m-xylene	2.979	2.878	3.078	50.220	50.000	0.4