

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM **Case No.:** P4368 **SAS No.:** P4368 **SDG NO.:** P4368

Instrument ID: ECD_D **Date(s) Analyzed:** 10/11/2024 10/11/2024

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.73	5.63	5.83	6501920
		2	6.31	6.21	6.41	23291500
		3	6.93	6.83	7.03	57507900
		4	7.02	6.92	7.12	41443300
		5	7.43	7.33	7.53	55926400

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Instrument ID: ECD_D Date(s) Analyzed: 10/11/2024 10/11/2024

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.13	5.03	5.23	56394300
		2	5.46	5.36	5.56	57154800
		3	5.82	5.72	5.92	51839100
		4	6.74	6.64	6.84	166014000
		5	7.19	7.09	7.29	110417000