

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM **Case No.:** P4368 **SAS No.:** P4368 **SDG NO.:** P4368

Instrument ID: ECD_D **Date(s) Analyzed:** 10/30/2024 10/30/2024

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.26	6.16	6.36	17010800
		2	6.46	6.36	6.56	24820600
		3	7.16	7.06	7.26	51666700
		4	7.58	7.48	7.68	60914100
		5	7.95	7.85	8.05	36712200

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM **Case No.:** P4368 **SAS No.:** P4368 **SDG NO.:** P4368

Instrument ID: ECD_D **Date(s) Analyzed:** 10/30/2024 10/30/2024

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.49	5.39	5.59	98589700
		2	5.66	5.56	5.76	65656500
		3	6.78	6.68	6.88	335187000
		4	7.22	7.12	7.32	238089000
		5	7.35	7.25	7.45	170925000