

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Continuing Calib Date: 10/17/2024 Initial Calibration Date(s): 10/11/2024 10/11/2024

Continuing Calib Time: 09:06 Initial Calibration Time(s): 11:15 12:09

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.90	8.90	8.80	9.00	0.00
Tetrachloro-m-xylene	3.44	3.44	3.34	3.54	0.00
alpha-BHC	3.89	3.89	3.79	3.99	0.00
beta-BHC	4.40	4.40	4.30	4.50	0.00
delta-BHC	4.64	4.64	4.54	4.74	0.00
gamma-BHC (Lindane)	4.21	4.21	4.11	4.31	0.00
Heptachlor	4.81	4.81	4.71	4.91	0.00
Aldrin	5.15	5.14	5.04	5.24	0.00
Heptachlor epoxide	5.56	5.56	5.46	5.66	0.00
Endosulfan I	5.95	5.94	5.84	6.04	-0.01
Dieldrin	6.22	6.22	6.12	6.32	0.00
4,4'-DDE	6.07	6.07	5.97	6.17	0.00
Endrin	6.44	6.44	6.34	6.54	0.00
Endosulfan II	6.66	6.66	6.56	6.76	0.00
4,4'-DDD	6.58	6.58	6.48	6.68	0.00
Endosulfan sulfate	7.02	7.02	6.92	7.12	0.00
4,4'-DDT	6.89	6.89	6.79	6.99	0.00
Methoxychlor	7.37	7.37	7.27	7.47	0.00
Endrin ketone	7.50	7.50	7.40	7.60	0.00
Endrin aldehyde	6.79	6.78	6.68	6.88	0.00
alpha-Chlordane	5.90	5.90	5.80	6.00	0.00
gamma-Chlordane	5.82	5.82	5.72	5.92	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Continuing Calib Date: 10/17/2024 Initial Calibration Date(s): 10/11/2024 10/11/2024

Continuing Calib Time: 09:06 Initial Calibration Time(s): 11:15 12:09

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
alpha-BHC	3.39	3.39	3.29	3.49	0.00
beta-BHC	4.02	4.02	3.92	4.12	0.01
delta-BHC	4.25	4.25	4.15	4.35	0.00
gamma-BHC (Lindane)	3.72	3.72	3.62	3.82	0.00
Heptachlor	4.07	4.07	3.97	4.17	0.00
Aldrin	4.36	4.36	4.26	4.46	0.00
Heptachlor epoxide	4.86	4.86	4.76	4.96	0.00
Endosulfan I	5.23	5.23	5.13	5.33	0.00
Dieldrin	5.50	5.50	5.40	5.60	0.00
4,4'-DDE	5.36	5.36	5.26	5.46	0.00
Endrin	5.78	5.78	5.68	5.88	0.00
Endosulfan II	6.07	6.07	5.97	6.17	0.00
4,4'-DDD	5.92	5.92	5.82	6.02	0.01
Endosulfan sulfate	6.47	6.47	6.37	6.57	0.00
4,4'-DDT	6.17	6.17	6.07	6.27	0.00
Methoxychlor	6.74	6.74	6.64	6.84	0.00
Endrin ketone	6.98	6.98	6.88	7.08	0.00
Endrin aldehyde	6.25	6.25	6.15	6.35	0.00
alpha-Chlordane	5.18	5.18	5.08	5.28	0.00
gamma-Chlordane	5.11	5.11	5.01	5.21	0.00