

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Continuing Calib Date: 10/17/2024 Initial Calibration Date(s): 10/11/2024 10/11/2024

Continuing Calib Time: 17:28 Initial Calibration Time(s): 11:15 12:09

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.90	8.90	8.80	9.00	0.00
Tetrachloro-m-xylene	3.44	3.44	3.34	3.54	0.00
alpha-BHC	3.89	3.89	3.79	3.99	0.01
delta-BHC	4.64	4.64	4.54	4.74	0.00
Aldrin	5.12	5.14	5.04	5.24	0.02
Heptachlor epoxide	5.56	5.56	5.46	5.66	0.00
Endosulfan I	5.96	5.94	5.84	6.04	-0.02
Dieldrin	6.21	6.22	6.12	6.32	0.01
Endrin	6.46	6.44	6.34	6.54	-0.02
Endosulfan II	6.68	6.66	6.56	6.76	-0.02
4,4'-DDD	6.57	6.58	6.48	6.68	0.01
Endosulfan sulfate	7.02	7.02	6.92	7.12	0.00
Endrin ketone	7.48	7.50	7.40	7.60	0.02
alpha-Chlordane	5.92	5.90	5.80	6.00	-0.02
gamma-Chlordane	5.84	5.82	5.72	5.92	-0.02

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Continuing Calib Date: 10/17/2024 Initial Calibration Date(s): 10/11/2024 10/11/2024

Continuing Calib Time: 17:28 Initial Calibration Time(s): 11:15 12:09

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
alpha-BHC	3.38	3.39	3.29	3.49	0.01
beta-BHC	4.03	4.02	3.92	4.12	-0.01
delta-BHC	4.25	4.25	4.15	4.35	0.00
gamma-BHC (Lindane)	3.72	3.72	3.62	3.82	0.00
Heptachlor	4.10	4.07	3.97	4.17	-0.03
Aldrin	4.36	4.36	4.26	4.46	0.00
Heptachlor epoxide	4.86	4.86	4.76	4.96	0.00
4,4'-DDE	5.35	5.36	5.26	5.46	0.01
Endrin	5.79	5.78	5.68	5.88	-0.01
Endosulfan II	6.07	6.07	5.97	6.17	0.00
4,4'-DDD	5.90	5.92	5.82	6.02	0.02
Endosulfan sulfate	6.48	6.47	6.37	6.57	-0.01
4,4'-DDT	6.18	6.17	6.07	6.27	-0.01
Methoxychlor	6.74	6.74	6.64	6.84	0.00
Endrin aldehyde	6.26	6.25	6.15	6.35	-0.01
alpha-Chlordane	5.20	5.18	5.08	5.28	-0.02
gamma-Chlordane	5.13	5.11	5.01	5.21	-0.02