

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/23/2024 09/23/2024

Client Sample No. (PEM): PEM - PL092320.D Date Analyzed: 10/11/2024

Lab Sample No.(PEM): PEM Time Analyzed: 08:56

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	9.059	8.960	9.160	22.700	20.000	13.5
Tetrachloro-m-xylene	3.540	3.490	3.590	22.460	20.000	12.3
alpha-BHC	3.995	3.940	4.050	11.450	10.000	14.5
beta-BHC	4.525	4.470	4.580	11.980	10.000	19.8
gamma-BHC (Lindane)	4.328	4.280	4.380	10.840	10.000	8.4
Endrin	6.575	6.500	6.650	45.920	50.000	-8.2
4,4'-DDT	7.025	6.950	7.100	96.120	100.000	-3.9
Methoxychlor	7.503	7.430	7.570	231.460	250.000	-7.4

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/23/2024 09/23/2024

Client Sample No. (PEM): PEM - PL092320.D Date Analyzed: 10/11/2024

Lab Sample No.(PEM): PEM Time Analyzed: 08:56

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	7.919	7.820	8.020	20.990	20.000	5.0
Tetrachloro-m-xylene	2.778	2.730	2.830	21.410	20.000	7.1
alpha-BHC	3.281	3.230	3.330	10.130	10.000	1.3
beta-BHC	3.911	3.860	3.960	11.680	10.000	16.8
gamma-BHC (Lindane)	3.611	3.560	3.660	9.770	10.000	-2.3
Endrin	5.643	5.570	5.710	48.200	50.000	-3.6
4,4'-DDT	6.042	5.970	6.110	108.430	100.000	8.4
Methoxychlor	6.617	6.550	6.690	243.010	250.000	-2.8