

## Report of Analysis

Client:	Chemtech Consulting Group		Date Collected:	10/09/24	
Project:	LOD-LOQ Study		Date Received:	10/09/24	
Client Sample ID:	LOD-MDL-WATER-01-QT4-2024		SDG No.:	P4368	
Lab Sample ID:	P4368-07		Matrix:	WATER	
Analytical Method:	SW8081		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	Pesticide-TCL	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD086517.D	1	10/10/24 11:57	11/11/24 13:16	PB164033

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
319-84-6	alpha-BHC	0.014	JP	0.0061	0.050	ug/L
319-85-7	beta-BHC	0.016	J	0.014	0.050	ug/L
319-86-8	delta-BHC	0.015	J	0.015	0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.014	JP	0.0049	0.050	ug/L
76-44-8	Heptachlor	0.015	JP	0.0054	0.050	ug/L
309-00-2	Aldrin	0.013	JP	0.0044	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.014	JP	0.0090	0.050	ug/L
959-98-8	Endosulfan I	0.014	J	0.0050	0.050	ug/L
60-57-1	Dieldrin	0.014	JP	0.0047	0.050	ug/L
72-55-9	4,4-DDE	0.014	JP	0.0045	0.050	ug/L
72-20-8	Endrin	0.016	JP	0.0043	0.050	ug/L
33213-65-9	Endosulfan II	0.015	J	0.0075	0.050	ug/L
72-54-8	4,4-DDD	0.014	JP	0.0092	0.050	ug/L
1031-07-8	Endosulfan Sulfate	0.015	JP	0.0035	0.050	ug/L
50-29-3	4,4-DDT	0.013	JP	0.0044	0.050	ug/L
72-43-5	Methoxychlor	0.014	J	0.011	0.050	ug/L
53494-70-5	Endrin ketone	0.016	JP	0.0097	0.050	ug/L
7421-93-4	Endrin aldehyde	0.014	J	0.0099	0.050	ug/L
5103-71-9	alpha-Chlordane	0.014	JP	0.0060	0.050	ug/L
5103-74-2	gamma-Chlordane	0.014	J	0.0060	0.050	ug/L
8001-35-2	Toxaphene	1.00	U	0.15	1.00	ug/L
<b>SURROGATES</b>						
2051-24-3	Decachlorobiphenyl	21.8		43 - 140	109%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.3		77 - 126	106%	SPK: 20

## Report of Analysis

Client:	Chemtech Consulting Group		Date Collected:	10/09/24	
Project:	LOD-LOQ Study		Date Received:	10/09/24	
Client Sample ID:	LOD-MDL-WATER-01-QT4-2024		SDG No.:	P4368	
Lab Sample ID:	P4368-07		Matrix:	WATER	
Analytical Method:	SW8081		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	Pesticide-TCL	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD086517.D	1	10/10/24 11:57	11/11/24 13:16	PB164033

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

### Comments:

U = Not Detected  
 LOQ = Limit of Quantitation  
 MDL = Method Detection Limit  
 LOD = Limit of Detection  
 E = Value Exceeds Calibration Range  
 P = Indicates >25% difference for detected concentrations between the two GC columns  
 Q = indicates LCS control criteria did not meet requirements  
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value  
 B = Analyte Found in Associated Method Blank  
 N = Presumptive Evidence of a Compound  
 \* = Values outside of QC limits  
 D = Dilution  
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.  
 () = Laboratory InHouse Limit