

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
 ProjectID: LOD-LOQ Study
 Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG No.: P4368

Calibration Sequence : FE101124		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	320763220	188684	FE050582.D
850	124012735	145897	FE050583.D
340	48708856	143261	FE050584.D
170	26268086	154518	FE050585.D
85	14743318	173451	FE050586.D
AVG RF : 161162		% RSD : 12.04	AVG RT : 14.9452