

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
 ProjectID: LOD-LOQ Study
 Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG No.: P4368

Calibration Sequence : FF100324		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	235707239	138651	FF014608.D
850	108720465	127906	FF014609.D
340	43597100	128227	FF014610.D
170	23322733	137193	FF014611.D
85	12811042	150718	FF014612.D
AVG RF : 136539		% RSD : 6.846	AVG RT : 14.8864