

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
 ProjectID: LOD-LOQ Study
 Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG No.: P4368

Calibration Sequence : FG100224		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	210997993	124116	FG014366.D
850	108163383	127251	FG014367.D
680	86368492	127012	FG014368.D
170	23504642	138263	FG014369.D
85	10806064	127130	FG014370.D
AVG RF : 128754		% RSD : 4.252	AVG RT : 14.915