

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM **Case No.:** P4368 **SAS No.:** P4368 **SDG NO.:** P4368

Instrument ID: ECD_R **Calibration Date(s):** 10/14/2024 10/14/2024
Calibration Times: 10:06 10:44

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 0.5 = <u>PR068662.D</u>		CF 0.25 = <u>PR068663.D</u>			
CF 0.1 = <u>PR068664.D</u>		CF 0.05 = <u>PR068665.D</u>		CF 0.025 = <u>PR068666.D</u>			
COMPOUND	CF 0.5	CF 0.25	CF 0.1	CF 0.05	CF 0.025	CF	% RSD
DBCP	25409600000	26804700000	28107800000	29145200000	25439100000	26981300000	6
EDB	13463800000	14256500000	14975600000	14804100000	15439800000	14588000000	5

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Instrument ID: ECD_R Calibration Date(s): 10/14/2024 10/14/2024

Calibration Times: 10:06 10:44

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 0.5 = <u>PR068662.D</u>		CF 0.25 = <u>PR068663.D</u>			
CF 0.1 = <u>PR068664.D</u>		CF 0.05 = <u>PR068665.D</u>		CF 0.025 = <u>PR068666.D</u>			
COMPOUND	CF 0.5	CF 0.25	CF 0.1	CF 0.05	CF 0.025	CF	% RSD
DBCP	132914000000	138198000000	144346000000	150368000000	128254000000	138816000000	6
EDB	63120300000	65967100000	74049900000	76936700000	65637500000	69142300000	9