

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM **Case No.:** P4368 **SAS No.:** P4368 **SDG NO.:** P4368

Continuing Calib Date: 10/14/2024 **Initial Calibration Date(s):** 10/14/2024 10/14/2024

Continuing Calib Time: 11:13 **Initial Calibration Time(s):** 10:06 10:44

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DBCP	6.58	6.58	6.48	6.68	0.00
EDB	2.84	2.83	2.73	2.93	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM **Case No.:** P4368 **SAS No.:** P4368 **SDG NO.:** P4368

Continuing Calib Date: 10/14/2024 **Initial Calibration Date(s):** 10/14/2024 10/14/2024

Continuing Calib Time: 11:13 **Initial Calibration Time(s):** 10:06 10:44

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DBCP	5.55	5.55	5.45	5.65	0.00
EDB	2.23	2.22	2.12	2.32	-0.01