

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM **Case No.:** P4368 **SAS No.:** P4368 **SDG NO.:** P4368

GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/14/2024 10/14/2024

Client Sample No.: CCAL01 **Date Analyzed:** 10/14/2024

Lab Sample No.: M8011-504 CCC 0. **Data File :** PR068669.D **Time Analyzed:** 11:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
DBCP	6.577	6.477	6.677	0.100	0.100	0.0
EDB	2.837	2.734	2.934	0.090	0.100	0.0

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/14/2024 10/14/2024

Client Sample No.: CCAL01 Date Analyzed: 10/14/2024

Lab Sample No.: M8011-504 CCC 0. Data File : PR068669.D Time Analyzed: 11:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
DBCP	5.549	5.449	5.649	0.100	0.100	0.0
EDB	2.226	2.122	2.322	0.100	0.100	0.0