

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/14/2024 10/14/2024

Client Sample No.: CCAL02 Date Analyzed: 10/14/2024

Lab Sample No.: M8011-504 CCC 0. Data File : PR068674.D Time Analyzed: 12:14

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
DBCP	6.575	6.477	6.677	0.100	0.100	0.0
EDB	2.836	2.734	2.934	0.090	0.100	0.0

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/14/2024 10/14/2024

Client Sample No.: CCAL02 Date Analyzed: 10/14/2024

Lab Sample No.: M8011-504 CCC 0. Data File : PR068674.D Time Analyzed: 12:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
DBCP	5.551	5.449	5.649	0.100	0.100	0.0
EDB	2.227	2.122	2.322	0.100	0.100	0.0