
From: Irving, Madison <Madison.Irving@gdit.com>
Sent: Thursday, October 24, 2024 10:34 AM
To: Sohil Jodhani; Nimisha Pandya; Mohammad Ahmed
Cc: R5 - R5RSCC@epa.gov (R5RSCC@epa.gov)
Subject: REGION 5 | CASE 51791 | LAB ACE | SDG E27G9 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT | FINAL
Attachments: E27G9.pdf; E27J1.pdf; Internal standard fail.pdf

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good morning ACE,

Regarding the laboratory's reported issue for Case 51791/SDG E27G9, please see the Region's input below and let me know if you have any questions.

LAB: "Lab has received water samples for SVOA-SIM analysis. Lab has performed undiluted SVOA-SIM analysis for the samples E27G9 & E27J1 and samples found positive with high concentration of target analytes detected and required dilution as you can see attached form-1 with quant report. Due to high concentration of target analytes and due to matrix interference, samples have one of the internal standard recoveries outside the QC limits therefore lab would like to confirm that lab will report undiluted SVOA-SIM analysis with internal standard failure and further dilution in final electronic deliverables.

Please see attachment for your reference."

REGION: "The lab's proposal is acceptable."

***** Please note that in order to waive/remove defects associated to this issue, please send a detailed waiver request to the appropriate ASB PM.

Thank you,

Madison Irving
Associate Environmental Analyst
GDIT Federal Civilian Division
Under Contract to EPA
T: 919-768-4149
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

LEAVE ALERT: None

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: R5RSCC <R5RSCC@epa.gov>
Sent: Wednesday, October 23, 2024 4:30 PM
To: Irving, Madison <Madison.Irving@gdit.com>
Subject: FW: REGION 5 | CASE 51791 | LAB ACE | SDG E27H4 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hi Madison,

The lab's proposal is acceptable.

Thanks,
Amanda

Amanda Wroble, Ph.D.
Metals Chemist/Group Leader & Regional Sample Control Coordinator (RSCC)/CRL Sample Coordinator
USEPA Region 5 LSASD/AIMB
Chicago Regional Laboratory
536 S. Clark St. LAB-10C, 10th Floor
Chicago, IL 60605
(312) 353-0375

From: Hack, Emily <ehack@eaest.com>
Sent: Wednesday, October 23, 2024 3:26 PM
To: R5RSCC <R5RSCC@epa.gov>
Cc: Bumba, Lauren <bumba.lauren@epa.gov>
Subject: RE: REGION 5 | CASE 51791 | LAB ACE | SDG E27H4 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Yes, that sounds acceptable.

Thanks,

Emily Hack
Data Scientist
Direct: (410) 329-5149

From: R5RSCC <R5RSCC@epa.gov>
Sent: Wednesday, October 23, 2024 4:23 PM
To: Hack, Emily <ehack@eaest.com>
Cc: Bumba, Lauren <bumba.lauren@epa.gov>
Subject: FW: REGION 5 | CASE 51791 | LAB ACE | SDG E27H4 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Hi Emily,

Is the lab's proposed plan of action below acceptable?

Thanks,
Amanda

Amanda Wroble, Ph.D.
Metals Chemist/Group Leader & Regional Sample Control Coordinator (RSCC)/CRL Sample Coordinator
USEPA Region 5 LSASD/AIMB
Chicago Regional Laboratory
536 S. Clark St. LAB-10C, 10th Floor
Chicago, IL 60605
(312) 353-0375

From: Irving, Madison <Madison.Irving@gdit.com>
Sent: Wednesday, October 23, 2024 10:27 AM
To: R5RSCC <R5RSCC@epa.gov>
Subject: REGION 5 | CASE 51791 | LAB ACE | SDG E27H4 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Good morning Region 5,

Please advise on the laboratory's reported issue below for Case 51791/SDG E27H4.

LAB: "Lab has received water samples for SVOA-SIM analysis. Lab has performed undiluted SVOA-SIM analysis for the samples E27G9 & E27J1 and samples found positive with high concentration of target analytes detected and required dilution as you can see attached form-1 with quant report. Due to high concentration of target analytes and due to matrix interference, samples have one of the internal standard recoveries outside the QC limits therefore lab would like to confirm that lab will report undiluted SVOA-SIM analysis with internal standard failure and further dilution in final electronic deliverables.

Please see attachment for your reference."

Thank you,

Madison Irving
Associate Environmental Analyst
GDIT Federal Civilian Division

Under Contract to EPA
T: 919-768-4149
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

LEAVE ALERT: None

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: DASsupport
Sent: Wednesday, October 23, 2024 9:40 AM
To: Sohil Jodhani <Sohil.Jodhani@alliancetg.com>
Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>
Subject: REGION 5 | CASE 51791 | LAB ACE | SDG E27H4 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Thank you for contacting DAS Support. This is to confirm that your email has been received and is currently being addressed. An updated response will be provided as soon as it is available.

Thank you,

The DAS Support Team
DASsupport@gdit.com

GDIT Federal Civilian Division
Under contract to the EPA
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Sohil Jodhani <Sohil.Jodhani@alliancetg.com>
Sent: Wednesday, October 23, 2024 9:01 AM

To: DASsupport <DASsupport@gdit.com>

Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>

Subject: REGION 5 | CASE 51791 | LAB ACE | SDG E27H4 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Good morning,

Lab is sending this email with regards to case 51791 and SDG E27H4.

Lab has received water samples for SVOA-SIM analysis. Lab has performed undiluted SVOA-SIM analysis for the samples E27G9 & E27J1 and samples found positive with high concentration of target analytes detected and required dilution as you can see attached form-1 with quant report. Due to high concentration of target analytes and due to matrix interference, samples have one of the internal standard recoveries outside the QC limits therefore lab would like to confirm that lab will report undiluted SVOA-SIM analysis with internal standard failure and further dilution in final electronic deliverables.

Please see attachment for your reference.

Thanks & Regards,



Sohil Jodhani

QA/QC Director

An Alliance Technical Group Company

Main: 908-789-8900

Direct: 908-728-3152

Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092

www.alliancetg.com



From: Irving, Madison <Madison.Irving@gdit.com>
Sent: Thursday, October 24, 2024 10:36 AM
To: Sohil Jodhani; Nimisha Pandya; Mohammad Ahmed
Cc: R5 - R5RSCC@epa. gov (R5RSCC@epa.gov)
Subject: REGION 5 | CASE 51791 | LAB ACE | SDG E27G9 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DASS SUPPORT | FINAL
Attachments: E27G8RX.pdf; E27H2RX.pdf; E27H5RX.pdf; E27H7RX.pdf; E27H7RXDL.pdf; E27J0RX.pdf

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good morning ACE,

Regarding the laboratory's reported issue for Case 51791/SDG E27G9, please see the Region's input below and let me know if you have any questions.

LAB: "Lab has received water samples for SVOA analysis. Lab has analyzed low level SVOA analysis for these samples and samples had two and/or more of the surrogates were not recovered and also samples were originally received with pH more than 12. As part of corrective action, Lab has re-extracted these samples for SVOA analysis and in re-extraction/re-analysis, samples have all surrogates recovered. However, re-extraction is done outside the holding time as Lab has only 5 days holding time for water samples from VTSR. In this case, lab will report both the data in final electronic deliverables."

Please see attachment for your reference.

Please see attachment for your reference."

REGION: "The lab's proposal is acceptable."

***** Please note that in order to waive/remove defects associated to this issue, please send a detailed waiver request to the appropriate ASB PM.

Thank you,

Madison Irving
Associate Environmental Analyst
GDIT Federal Civilian Division
Under Contract to EPA
T: 919-768-4149
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

LEAVE ALERT: None

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: R5RSCC <R5RSCC@epa.gov>
Sent: Wednesday, October 23, 2024 4:30 PM
To: Irving, Madison <Madison.Irving@gdit.com>
Subject: FW: REGION 5 | CASE 51791 | LAB ACE | SDG E27G9 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DASS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hi Madison,

The lab's proposal is acceptable.

Thanks,
Amanda

Amanda Wroble, Ph.D.
Metals Chemist/Group Leader & Regional Sample Control Coordinator (RSCC)/CRL Sample Coordinator
USEPA Region 5 LSASD/AIMB
Chicago Regional Laboratory
536 S. Clark St. LAB-10C, 10th Floor
Chicago, IL 60605
(312) 353-0375

From: Hack, Emily <ehack@eaest.com>
Sent: Wednesday, October 23, 2024 3:28 PM
To: R5RSCC <R5RSCC@epa.gov>
Cc: Bumba, Lauren <bumba.lauren@epa.gov>
Subject: RE: REGION 5 | CASE 51791 | LAB ACE | SDG E27G9 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DASS SUPPORT

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Yes, that sounds acceptable.

Emily Hack
Data Scientist
Direct: (410) 329-5149

From: R5RSCC <R5RSCC@epa.gov>
Sent: Wednesday, October 23, 2024 4:25 PM
To: Hack, Emily <ehack@eaest.com>
Cc: Bumba, Lauren <bumba.lauren@epa.gov>
Subject: FW: REGION 5 | CASE 51791 | LAB ACE | SDG E27G9 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DASS SUPPORT

Hi Emily,

Is the lab's proposal below acceptable?

Thanks,
Amanda

Amanda Wroble, Ph.D.
Metals Chemist/Group Leader & Regional Sample Control Coordinator (RSCC)/CRL Sample Coordinator
USEPA Region 5 LSASD/AIMB
Chicago Regional Laboratory
536 S. Clark St. LAB-10C, 10th Floor
Chicago, IL 60605
(312) 353-0375

From: Irving, Madison <Madison.Irving@gdit.com>
Sent: Wednesday, October 23, 2024 10:27 AM
To: R5RSCC <R5RSCC@epa.gov>
Subject: REGION 5 | CASE 51791 | LAB ACE | SDG E27G9 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DASS SUPPORT

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Good morning Region 5,

Please advise on the laboratory's reported issue below for Case 51791/SDG E27G9.

LAB: "Lab has received water samples for SVOA analysis. Lab has analyzed low level SVOA analysis for these samples and samples had two and/or more of the surrogates were not recovered and also samples were originally received with pH more than 12. As part of corrective action, Lab has re-extracted these samples for SVOA analysis and in re-extraction/re-analysis, samples have all surrogates recovered. However, re-extraction is done outside the holding time as Lab has only 5 days holding time for water samples from VTSR. In this case, lab will report both the data in final electronic deliverables.

Please see attachment for your reference."

Thank you,

Madison Irving
Associate Environmental Analyst
GDIT Federal Civilian Division
Under Contract to EPA
T: 919-768-4149
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com



LEAVE ALERT: None

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: DASSupport
Sent: Wednesday, October 23, 2024 9:40 AM
To: Sohil Jodhani <Sohil.Jodhani@alliancetg.com>
Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>
Subject: REGION 5 | CASE 51791 | LAB ACE | SDG E27G9 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DASS SUPPORT

Thank you for contacting DAS Support. This is to confirm that your email has been received and is currently being addressed. An updated response will be provided as soon as it is available.

Thank you,

The DAS Support Team
DASSupport@gdit.com

GDIT Federal Civilian Division
Under contract to the EPA
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com



This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Sohil Jodhani <Sohil.Jodhani@alliancetg.com>

Sent: Wednesday, October 23, 2024 9:22 AM

To: DASSupport <DASSupport@gdit.com>

Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>

Subject: REGION 5 | CASE 51791 | LAB ACE | SDG E27G9 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DASS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hello,

Lab is sending this email with reference to case 51791 and SDG E27G9.

Lab has received water samples for SVOA analysis. Lab has analyzed low level SVOA analysis for these samples and samples had two and/or more of the surrogates were not recovered and also samples were originally received with pH more than 12. As part of corrective action, Lab has re-extracted these samples for SVOA analysis and in re-extraction/re-analysis, samples have all surrogates recovered. However, re-extraction is done outside the holding time as Lab has only 5 days holding time for water samples from VTSR. In this case, lab will report both the data in final electronic deliverables.

Please see attachment for your reference.

Thanks & Regards,



Sohil Jodhani

QA/QC Director

An Alliance Technical Group Company

Main: 908-789-8900

Direct: 908-728-3152

Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092

www.alliancetg.com



From: Zafar, Tasmia (NE) <Tasmia.Zafar@gdit.com>
Sent: Tuesday, October 15, 2024 3:36 PM
To: Sohil Jodhani; Deepak Parmar; Mohammad Ahmed
Cc: R5RSCC
Subject: Region 05 | Case 51791 | Lab ACE | Issue Discrepancies with tags, jars, and/or COC | FINAL
Attachments: SKM_95824101010190.pdf

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good afternoon,

Issue: The laboratory received water samples for Case 51791 and some samples are listed with only SVOA analysis, but both SVOA and SVOA SIM analysis are scheduled. The laboratory would like direction on how to proceed.

Resolution: Per Region 5, the laboratory will note the issue in the SDG Narrative and proceed with the analysis of the samples for SVOA as listed on the COC. This resolution can be applied to all samples for this Case.

Please note that the laboratory may contact the appropriate CLP PM should any defects need to be waived for this issue.

Best Regards,
Tasmia Zafar
Associate Environmental Analyst
CLP QSS Coordinator – EPA Regions 5 & 6

T: (919) 768-4086
tasmia.zafar@gdit.com
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

Leave Alert: None

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: R5RSCC <R5RSCC@epa.gov>
Sent: Tuesday, October 15, 2024 3:30 PM
To: Zafar, Tasmia (NE) <Tasmia.Zafar@gdit.com>
Subject: RE: Region 05 | Case 51791 | Lab ACE | Issue Discrepancies with tags, jars, and/or COC

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hi Tasmia,

Yes please apply to all samples for this case.

Thanks,
Amanda

Amanda Wroble, Ph.D.
Metals Chemist/Group Leader & Regional Sample Control Coordinator (RSCC)/CRL Sample Coordinator
USEPA Region 5 LSASD/AIMB
Chicago Regional Laboratory
536 S. Clark St. LAB-10C, 10th Floor
Chicago, IL 60605
(312) 353-0375

From: Zafar, Tasmia (NE) <Tasmia.Zafar@gdit.com>

Sent: Tuesday, October 15, 2024 2:28 PM

To: R5RSCC <R5RSCC@epa.gov>

Subject: RE: Region 05 | Case 51791 | Lab ACE | Issue Discrepancies with tags, jars, and/or COC

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Hi Amanda,

Can this resolution be applied to all samples for this Case?

Best Regards,
Tasmia Zafar
Associate Environmental Analyst
CLP QSS Coordinator – EPA Regions 5 & 6

T: (919) 768-4086
tasmia.zafar@gdit.com
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

Leave Alert: None

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly

prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: R5RSCC <R5RSCC@epa.gov>
Sent: Tuesday, October 15, 2024 2:21 PM
To: Zafar, Tasmia (NE) <Tasmia.Zafar@gdit.com>
Subject: RE: Region 05 | Case 51791 | Lab ACE | Issue Discrepancies with tags, jars, and/or COC

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hi Tasmia,

Per the client, please have the lab proceed per the COCs. The samples with SVOA only are the blanks,

Thanks,
Amanda

Amanda Wroble, Ph.D.
Metals Chemist/Group Leader & Regional Sample Control Coordinator (RSCC)/CRL Sample Coordinator
USEPA Region 5 LSASD/AIMB
Chicago Regional Laboratory
536 S. Clark St. LAB-10C, 10th Floor
Chicago, IL 60605
(312) 353-0375

From: Zafar, Tasmia (NE) <Tasmia.Zafar@gdit.com>
Sent: Tuesday, October 15, 2024 11:59 AM
To: R5RSCC <R5RSCC@epa.gov>
Subject: RE: Region 05 | Case 51791 | Lab ACE | Issue Discrepancies with tags, jars, and/or COC

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Good afternoon,

I would like to follow up on the below.

Best Regards,
Tasmia Zafar
Associate Environmental Analyst
CLP QSS Coordinator – EPA Regions 5 & 6

T: (919) 768-4086
tasmia.zafar@gdit.com
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

Leave Alert: None

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Zafar, Tasmia (NE)
Sent: Friday, October 11, 2024 3:20 PM
To: R5RSCC <R5RSCC@epa.gov>
Subject: RE: Region 05 | Case 51791 | Lab ACE | Issue Discrepancies with tags, jars, and/or COC

Good afternoon,

I would like follow up on the below.

Best Regards,
Tasmia Zafar
Associate Environmental Analyst
CLP QSS Coordinator – EPA Regions 5 & 6

T: (919) 768-4086
tasmia.zafar@gdit.com
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

Leave Alert: None

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Zafar, Tasmia (NE)
Sent: Thursday, October 10, 2024 1:11 PM
To: R5RSCC <R5RSCC@epa.gov>
Subject: Region 05 | Case 51791 | Lab ACE | Issue Discrepancies with tags, jars, and/or COC

Good afternoon,

Please see the below issue.

Issue: The laboratory received water samples for Case 51791 and some samples are listed with only SVOA analysis, but both SVOA and SVOA SIM analysis are scheduled. The laboratory would like direction on how to proceed.

Best Regards,
Tasmia Zafar
Associate Environmental Analyst
CLP QSS Coordinator – EPA Regions 5 & 6

T: (919) 768-4086
tasmia.zafar@gdit.com
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

Leave Alert: None

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Deepak Parmar <Deepak.Parmar@alliancetg.com>
Sent: Thursday, October 10, 2024 12:30 PM
To: Zafar, Tasmia (NE) <Tasmia.Zafar@gdit.com>
Subject: Region 05 | Case 51791 | Lab CHM | Issue Discrepancies with tags, jars, and/or COC

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Good afternoon,

Issue: The laboratory received water samples for case 51791. Few sample on COC did not mentioned SVOC-SIM analysis as per ASR SOC-SIM analysis is scheduled . there for The laboratory would like to know how to proceed ?

Please see attachment for your reference.

Thanks & Regards,



Deepak Parmar
QA/QC
An Alliance Technical Group Company
Main: 908-789-8900
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com

