

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132883

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/11/2024	08:50
2	CAL2	1	Water	NA	NA	20.2	7.00	10/11/2024	08:51
3	CAL3	1	Water	NA	NA	20.2	10.02	10/11/2024	08:53
4	ICV	1	Water	NA	NA	20.3	7.00	10/11/2024	08:55
5	CCV1	1	Water	NA	NA	20.3	2.01	10/11/2024	09:00
6	P4385-02	1	Solid	20.02	20	20.6	4.96	10/11/2024	09:10
7	P4385-02DUP	1	Solid	20.03	20	20.7	4.97	10/11/2024	09:11
8	P4385-04	1	Solid	20.02	20	20.8	5.18	10/11/2024	09:15
9	P4385-06	1	Solid	20.01	20	20.3	5.29	10/11/2024	09:18
10	P4385-08	1	Solid	20.03	20	20.4	5.23	10/11/2024	09:20
11	P4385-10	1	Solid	20.03	20	20.5	6.43	10/11/2024	09:25
12	P4385-12	1	Solid	20.04	20	20.5	6.24	10/11/2024	09:30
13	P4385-14	1	Solid	20.02	20	20.7	6.84	10/11/2024	09:33
14	P4385-16	1	Solid	20.03	20	20.8	7.08	10/11/2024	09:36
15	P4385-18	1	Solid	20.04	20	20.4	7.12	10/11/2024	09:44
16	CCV2	1	Water	NA	NA	20.3	12.02	10/11/2024	09:47
17	P4385-20	1	Solid	20.02	20	20.5	7.19	10/11/2024	09:50
18	CCV3	1	Water	NA	NA	20.2	2.01	10/11/2024	09:51

WORKLIST(Hardcopy Internal Chain)

VB 132883

WorkList Name : ph p4385 WorkList ID : 184347 Department : Wet-Chemistry Date : 10-11-2024 08:39:36

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4385-02	SP-1	Solid	pH	Cool 4 deg C	SCHE03	K51	10/10/2024	9045D
P4385-04	SP-2	Solid	pH	Cool 4 deg C	SCHE03	K51	10/10/2024	9045D
P4385-06	SP-3	Solid	pH	Cool 4 deg C	SCHE03	K51	10/10/2024	9045D
P4385-08	SP-4	Solid	pH	Cool 4 deg C	SCHE03	K51	10/10/2024	9045D
P4385-10	SP-5	Solid	pH	Cool 4 deg C	SCHE03	K51	10/10/2024	9045D
P4385-12	SP-6	Solid	pH	Cool 4 deg C	SCHE03	K51	10/10/2024	9045D
P4385-14	SP-7	Solid	pH	Cool 4 deg C	SCHE03	K51	10/10/2024	9045D
P4385-16	SP-8	Solid	pH	Cool 4 deg C	SCHE03	K51	10/10/2024	9045D
P4385-18	SP-9	Solid	pH	Cool 4 deg C	SCHE03	K51	10/10/2024	9045D
P4385-20	SP-10	Solid	pH	Cool 4 deg C	SCHE03	K51	10/10/2024	9045D

Date/Time 10-11-24 08:45
 Raw Sample Received by: JH G201
 Raw Sample Relinquished by: JH G201

Date/Time 10-11-24 11:40
 Raw Sample Received by: JH G201
 Raw Sample Relinquished by: JH G201