



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Scheideler Excavating Co. Inc.

SDG No.: P4385

Project: Robbinsville

RunNo.: LB132883

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.00	7	100	90-110	10/11/2024
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	10/11/2024
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	10/11/2024
Sample ID: CCV3 pH	pH	2.01	2.00	101	90-110	10/11/2024

Initial and Continuing Calibration Verification

Client: Scheideler Excavating Co. Inc.

SDG No.: P4385

Project: Robbinsville

RunNo.: LB132887

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.502	0.5	100	90-110	10/11/2024
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	10/11/2024
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.503	0.5	101	90-110	10/11/2024
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	10/11/2024

Initial and Continuing Calibration Verification

Client: Scheideler Excavating Co. Inc.

SDG No.: P4385

Project: Robbinsville

RunNo.: LB132925

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.1	0.099	101	90-110	10/14/2024
Sample ID: CCV1 Cyanide	mg/L	0.25	0.25	100	90-110	10/14/2024
Sample ID: CCV2 Cyanide	mg/L	0.26	0.25	104	90-110	10/14/2024
Sample ID: CCV3 Cyanide	mg/L	0.25	0.25	100	90-110	10/14/2024
Sample ID: CCV4 Cyanide	mg/L	0.26	0.25	104	90-110	10/14/2024