

Laboratory Control Sample Summary

Client: Scheideler Excavating Co. Inc.

SDG No.: P4385

Project: Robbinsville

Run No.: LB132887

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB164052BS							
Hexavalent Chromium	mg/Kg	20	20.0		100	1	84-110	10/11/2024

Laboratory Control Sample Summary

Client: Scheideler Excavating Co. Inc.

SDG No.: P4385

Project: Robbinsville

Run No.: LB132925

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB164122BS							
Cyanide	mg/Kg	5	5.00		100	1	85-115	10/14/2024