

Continuing Calibration Report for SequenceID : FC101124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067399.D

Aliphatic C9-C12	7575097.000	60.000	3.175	6.457	126251.617	128354.488	1.638
Aliphatic C12-C16	4987010.000	40.000	6.458	9.848	124675.250	130957.677	4.797
Aliphatic C16-C21	7246870.000	60.000	9.849	13.207	120781.167	131476.102	8.135
Aliphatic C21-C28	9150736.000	80.000	13.208	16.862	114384.200	126823.664	9.808
Aliphatic C28-C40	11671078.000	120.000	16.863	21.712	97258.983	105215.848	7.562
Aliphatic EPH	40630791.000	360.000	3.175	21.712	112863.308	121110.937	6.810