

Continuing Calibration Report for SequenceID : FC101124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067408.D

Aliphatic C9-C12	7676829.000	60.000	3.175	6.457	127947.150	128354.488	0.317
Aliphatic C12-C16	4998555.000	40.000	6.458	9.848	124963.875	130957.677	4.577
Aliphatic C16-C21	7249603.000	60.000	9.849	13.207	120826.717	131476.102	8.100
Aliphatic C21-C28	9222288.000	80.000	13.208	16.862	115278.600	126823.664	9.103
Aliphatic C28-C40	12019967.000	120.000	16.863	21.712	100166.392	105215.848	4.799
Aliphatic EPH	41167242.000	360.000	3.175	21.712	114353.450	121110.937	5.580