

CALIBRATION VERIFICATION SUMMARY

Contract: SCHE03

Lab Code: CHEM Case No.: P4385 SAS No.: P4385 SDG NO.: P4385

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 10/07/2024 10/07/2024

Client Sample No.: CCAL01 Date Analyzed: 10/11/2024

Lab Sample No.: HSTDCCC750 Data File : PS027905.D Time Analyzed: 10:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.318	9.223	9.423	719.450	712.500	1.0
2,4,5-TP(Silvex)	9.034	8.938	9.138	717.440	712.500	0.7
2,4-D	8.181	8.085	8.285	686.600	705.000	-2.6
2,4-DB	9.881	9.785	9.985	702.270	712.500	-1.4
2,4-DCAA	7.095	6.998	7.198	735.100	750.000	-2.0
DICAMBA	7.274	7.177	7.377	705.590	705.000	0.1
DICHLORPROP	7.959	7.863	8.063	695.860	705.000	-1.3
Dinoseb	11.057	10.961	11.161	707.700	705.000	0.4

CALIBRATION VERIFICATION SUMMARY

Contract: SCHE03

Lab Code: CHEM Case No.: P4385 SAS No.: P4385 SDG NO.: P4385

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/07/2024 10/07/2024

Client Sample No.: CCAL01 Date Analyzed: 10/11/2024

Lab Sample No.: HSTDCCC750 Data File : PS027905.D Time Analyzed: 10:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.153	10.058	10.258	708.250	712.500	-0.6
2,4,5-TP(Silvex)	9.739	9.643	9.843	721.720	712.500	1.3
2,4-D	8.849	8.752	8.952	672.380	705.000	-4.6
2,4-DB	10.714	10.619	10.819	705.100	712.500	-1.0
2,4-DCAA	7.622	7.528	7.728	698.610	750.000	-6.9
DICAMBA	7.817	7.722	7.922	732.660	705.000	3.9
DICHLORPROP	8.524	8.428	8.628	690.830	705.000	-2.0
Dinoseb	11.088	10.993	11.193	705.540	705.000	0.1