

CALIBRATION VERIFICATION SUMMARY

Contract: SCHE03

Lab Code: CHEM Case No.: P4385 SAS No.: P4385 SDG NO.: P4385

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 10/07/2024 10/07/2024

Client Sample No.: CCAL03 Date Analyzed: 10/14/2024

Lab Sample No.: HSTDCCC750 Data File : PS027919.D Time Analyzed: 13:17

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.325	9.223	9.423	779.050	712.500	9.3
2,4,5-TP(Silvex)	9.040	8.938	9.138	781.510	712.500	9.7
2,4-D	8.187	8.085	8.285	750.840	705.000	6.5
2,4-DB	9.888	9.785	9.985	764.330	712.500	7.3
2,4-DCAA	7.099	6.998	7.198	794.520	750.000	5.9
DICAMBA	7.278	7.177	7.377	769.850	705.000	9.2
DICHLORPROP	7.964	7.863	8.063	748.400	705.000	6.2
Dinoseb	11.065	10.961	11.161	783.380	705.000	11.1

CALIBRATION VERIFICATION SUMMARY

Contract: SCHE03

Lab Code: CHEM Case No.: P4385 SAS No.: P4385 SDG NO.: P4385

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/07/2024 10/07/2024

Client Sample No.: CCAL03 Date Analyzed: 10/14/2024

Lab Sample No.: HSTDCCC750 Data File : PS027919.D Time Analyzed: 13:17

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.151	10.058	10.258	754.110	712.500	5.8
2,4,5-TP(Silvex)	9.736	9.643	9.843	778.620	712.500	9.3
2,4-D	8.845	8.752	8.952	760.550	705.000	7.9
2,4-DB	10.713	10.619	10.819	747.140	712.500	4.9
2,4-DCAA	7.622	7.528	7.728	784.440	750.000	4.6
DICAMBA	7.816	7.722	7.922	772.700	705.000	9.6
DICHLORPROP	8.521	8.428	8.628	724.490	705.000	2.8
Dinoseb	11.089	10.993	11.193	723.760	705.000	2.7