

CALIBRATION VERIFICATION SUMMARY

Contract: SCHE03

Lab Code: CHEM Case No.: P4385 SAS No.: P4385 SDG NO.: P4385

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 10/07/2024 10/07/2024

Client Sample No.: CCAL04 Date Analyzed: 10/14/2024

Lab Sample No.: HSTDCCC750 Data File : PS027930.D Time Analyzed: 20:13

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.319	9.223	9.423	765.640	712.500	7.5
2,4,5-TP(Silvex)	9.035	8.938	9.138	762.320	712.500	7.0
2,4-D	8.182	8.085	8.285	738.140	705.000	4.7
2,4-DB	9.883	9.785	9.985	746.440	712.500	4.8
2,4-DCAA	7.096	6.998	7.198	783.520	750.000	4.5
DICAMBA	7.275	7.177	7.377	751.460	705.000	6.6
DICHLORPROP	7.960	7.863	8.063	743.090	705.000	5.4
Dinoseb	11.058	10.961	11.161	758.640	705.000	7.6

CALIBRATION VERIFICATION SUMMARY

Contract: SCHE03

Lab Code: CHEM Case No.: P4385 SAS No.: P4385 SDG NO.: P4385

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/07/2024 10/07/2024

Client Sample No.: CCAL04 Date Analyzed: 10/14/2024

Lab Sample No.: HSTDCCC750 Data File : PS027930.D Time Analyzed: 20:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.153	10.058	10.258	706.040	712.500	-0.9
2,4,5-TP(Silvex)	9.738	9.643	9.843	690.700	712.500	-3.1
2,4-D	8.848	8.752	8.952	682.210	705.000	-3.2
2,4-DB	10.715	10.619	10.819	668.500	712.500	-6.2
2,4-DCAA	7.624	7.528	7.728	720.250	750.000	-4.0
DICAMBA	7.818	7.722	7.922	691.530	705.000	-1.9
DICHLORPROP	8.523	8.428	8.628	665.930	705.000	-5.5
Dinoseb	11.088	10.993	11.193	686.300	705.000	-2.7