

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4385

Client: Scheideler Excavating Co. Inc.

Analytical Method: 8151A

DataFile : PS027910.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: P4385-02MS	SP-1MS											
	DICAMBA	178	0	94.8	ug/Kg	53	*			70 (10)	130 (112)	
	DICHLORPROP	178	0	80.0	ug/Kg	45	*			70 (10)	130 (113)	
	2,4-D	178	0	102	ug/Kg	57	*			70 (10)	130 (144)	
	2,4,5-TP(Silvex)	178	0	56.7	ug/Kg	32	*			70 (10)	130 (114)	
	2,4,5-T	178	0	72.3	ug/Kg	41	*			70 (10)	130 (115)	
	2,4-DB	178	0	88.2	ug/Kg	50	*			70 (10)	130 (140)	
	Dinoseb	178	0	0	ug/Kg	0	*			70 (10)	130 (118)	