

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4385

Client: Scheideler Excavating Co. Inc.

Analytical Method: 8151A

DataFile : PS027911.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: P4385-02MSD	SP-1MSD											
	DICAMBA	178.1	0	96.3	ug/Kg	54	*	2		70 (10)	130 (112)	30 (20)
	DICHLORPROP	178.1	0	81.1	ug/Kg	46	*	2		70 (10)	130 (113)	30 (20)
	2,4-D	178.1	0	101	ug/Kg	57	*	0		70 (10)	130 (144)	30 (20)
	2,4,5-TP(Silvex)	178.1	0	55.6	ug/Kg	31	*	3		70 (10)	130 (114)	30 (20)
	2,4,5-T	178.1	0	71.4	ug/Kg	40	*	2		70 (10)	130 (115)	30 (20)
	2,4-DB	178.1	0	55.3	ug/Kg	31	*	47	*	70 (10)	130 (140)	30 (20)
	Dinoseb	178.1	0	0	ug/Kg	0	*	0		70 (10)	130 (118)	30 (20)