

CALIBRATION VERIFICATION SUMMARY

Contract: SCHE03

Lab Code: CHEM Case No.: P4385 SAS No.: P4385 SDG NO.: P4385

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL01 Date Analyzed: 10/11/2024

Lab Sample No.: AR1660CCC500 Data File : PP067740.D Time Analyzed: 09:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.918	5.819	6.019	509.520	500.000	1.9
Aroclor-1016-2	5.941	5.842	6.042	496.280	500.000	-0.7
Aroclor-1016-3	6.005	5.905	6.105	501.510	500.000	0.3
Aroclor-1016-4	6.103	6.003	6.203	493.640	500.000	-1.3
Aroclor-1016-5	6.397	6.298	6.498	487.460	500.000	-2.5
Aroclor-1260-1	7.521	7.422	7.622	466.340	500.000	-6.7
Aroclor-1260-2	7.774	7.675	7.875	471.010	500.000	-5.8
Aroclor-1260-3	8.135	8.037	8.237	459.690	500.000	-8.1
Aroclor-1260-4	8.374	8.275	8.475	470.660	500.000	-5.9
Aroclor-1260-5	8.711	8.612	8.812	483.490	500.000	-3.3
Decachlorobiphenyl	10.668	10.569	10.769	50.020	50.000	0.0
Tetrachloro-m-xylene	4.754	4.655	4.855	50.420	50.000	0.8

CALIBRATION VERIFICATION SUMMARY

Contract: SCHE03

Lab Code: CHEM Case No.: P4385 SAS No.: P4385 SDG NO.: P4385

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL01 Date Analyzed: 10/11/2024

Lab Sample No.: AR1660CCC500 Data File : PP067740.D Time Analyzed: 09:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.160	5.061	5.261	476.060	500.000	-4.8
Aroclor-1016-2	5.179	5.081	5.281	480.500	500.000	-3.9
Aroclor-1016-3	5.360	5.262	5.462	489.680	500.000	-2.1
Aroclor-1016-4	5.399	5.301	5.501	486.210	500.000	-2.8
Aroclor-1016-5	5.618	5.520	5.720	482.760	500.000	-3.4
Aroclor-1260-1	6.662	6.564	6.764	465.490	500.000	-6.9
Aroclor-1260-2	6.848	6.750	6.950	470.020	500.000	-6.0
Aroclor-1260-3	7.006	6.908	7.108	463.750	500.000	-7.3
Aroclor-1260-4	7.481	7.383	7.583	463.740	500.000	-7.3
Aroclor-1260-5	7.719	7.622	7.822	468.920	500.000	-6.2
Decachlorobiphenyl	9.214	9.119	9.319	46.980	50.000	-6.0
Tetrachloro-m-xylene	4.051	3.952	4.152	48.260	50.000	-3.5