

CALIBRATION VERIFICATION SUMMARY

Contract: SCHE03

Lab Code: CHEM Case No.: P4385 SAS No.: P4385 SDG NO.: P4385

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL02 Date Analyzed: 10/11/2024

Lab Sample No.: AR1660CCC500 Data File : PP067756.D Time Analyzed: 16:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.918	5.819	6.019	499.260	500.000	-0.1
Aroclor-1016-2	5.940	5.842	6.042	491.190	500.000	-1.8
Aroclor-1016-3	6.004	5.905	6.105	491.960	500.000	-1.6
Aroclor-1016-4	6.102	6.003	6.203	491.910	500.000	-1.6
Aroclor-1016-5	6.395	6.298	6.498	497.590	500.000	-0.5
Aroclor-1260-1	7.520	7.422	7.622	477.650	500.000	-4.5
Aroclor-1260-2	7.773	7.675	7.875	477.550	500.000	-4.5
Aroclor-1260-3	8.134	8.037	8.237	479.780	500.000	-4.0
Aroclor-1260-4	8.371	8.275	8.475	475.540	500.000	-4.9
Aroclor-1260-5	8.709	8.612	8.812	478.090	500.000	-4.4
Decachlorobiphenyl	10.664	10.569	10.769	49.000	50.000	-2.0
Tetrachloro-m-xylene	4.754	4.655	4.855	49.740	50.000	-0.5

CALIBRATION VERIFICATION SUMMARY

Contract: SCHE03

Lab Code: CHEM Case No.: P4385 SAS No.: P4385 SDG NO.: P4385

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL02 Date Analyzed: 10/11/2024

Lab Sample No.: AR1660CCC500 Data File : PP067756.D Time Analyzed: 16:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.159	5.061	5.261	478.370	500.000	-4.3
Aroclor-1016-2	5.180	5.081	5.281	478.910	500.000	-4.2
Aroclor-1016-3	5.360	5.262	5.462	491.450	500.000	-1.7
Aroclor-1016-4	5.399	5.301	5.501	487.170	500.000	-2.6
Aroclor-1016-5	5.618	5.520	5.720	499.380	500.000	-0.1
Aroclor-1260-1	6.662	6.564	6.764	475.180	500.000	-5.0
Aroclor-1260-2	6.849	6.750	6.950	474.650	500.000	-5.1
Aroclor-1260-3	7.007	6.908	7.108	478.930	500.000	-4.2
Aroclor-1260-4	7.481	7.383	7.583	475.530	500.000	-4.9
Aroclor-1260-5	7.720	7.622	7.822	468.280	500.000	-6.3
Decachlorobiphenyl	9.215	9.119	9.319	47.250	50.000	-5.5
Tetrachloro-m-xylene	4.051	3.952	4.152	48.980	50.000	-2.0