

CALIBRATION VERIFICATION SUMMARY

Contract: SCHE03

Lab Code: CHEM Case No.: P4385 SAS No.: P4385 SDG NO.: P4385

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL03 Date Analyzed: 10/11/2024

Lab Sample No.: AR1660CCC500 Data File : PP067771.D Time Analyzed: 20:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.917	5.819	6.019	501.520	500.000	0.3
Aroclor-1016-2	5.939	5.842	6.042	495.300	500.000	-0.9
Aroclor-1016-3	6.003	5.905	6.105	498.980	500.000	-0.2
Aroclor-1016-4	6.101	6.003	6.203	491.030	500.000	-1.8
Aroclor-1016-5	6.395	6.298	6.498	497.020	500.000	-0.6
Aroclor-1260-1	7.519	7.422	7.622	477.900	500.000	-4.4
Aroclor-1260-2	7.772	7.675	7.875	473.830	500.000	-5.2
Aroclor-1260-3	8.134	8.037	8.237	477.770	500.000	-4.4
Aroclor-1260-4	8.372	8.275	8.475	472.560	500.000	-5.5
Aroclor-1260-5	8.708	8.612	8.812	477.560	500.000	-4.5
Decachlorobiphenyl	10.664	10.569	10.769	49.130	50.000	-1.7
Tetrachloro-m-xylene	4.753	4.655	4.855	50.380	50.000	0.8

CALIBRATION VERIFICATION SUMMARY

Contract: SCHE03

Lab Code: CHEM Case No.: P4385 SAS No.: P4385 SDG NO.: P4385

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL03 Date Analyzed: 10/11/2024

Lab Sample No.: AR1660CCC500 Data File : PP067771.D Time Analyzed: 20:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.160	5.061	5.261	485.890	500.000	-2.8
Aroclor-1016-2	5.180	5.081	5.281	487.420	500.000	-2.5
Aroclor-1016-3	5.360	5.262	5.462	491.980	500.000	-1.6
Aroclor-1016-4	5.399	5.301	5.501	488.190	500.000	-2.4
Aroclor-1016-5	5.618	5.520	5.720	490.540	500.000	-1.9
Aroclor-1260-1	6.662	6.564	6.764	472.260	500.000	-5.5
Aroclor-1260-2	6.849	6.750	6.950	473.030	500.000	-5.4
Aroclor-1260-3	7.006	6.908	7.108	469.950	500.000	-6.0
Aroclor-1260-4	7.482	7.383	7.583	464.260	500.000	-7.1
Aroclor-1260-5	7.720	7.622	7.822	462.070	500.000	-7.6
Decachlorobiphenyl	9.215	9.119	9.319	46.350	50.000	-7.3
Tetrachloro-m-xylene	4.051	3.952	4.152	49.760	50.000	-0.5