

Surrogate Summary

SDG No.: **P4385**

Client: **Scheideler Excavating Co. Inc.**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP067586.D	PIBLK-PP067586.D	Tetrachloro-m-xylene	1	20	21.1	105		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	23.6	118		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	22.1	110		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	23.5	117		70 (60)	130 (140)
I.BLK-PP067744.D	PIBLK-PP067744.D	Tetrachloro-m-xylene	1	20	21.6	108		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	23.6	118		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	20.6	103		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	22.6	113		70 (60)	130 (140)
PB164066BL	PB164066BL	Tetrachloro-m-xylene	1	20	19.2	96		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	20.0	100		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.0	95		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	20.1	101		30 (32)	150 (175)
PB164066BS	PB164066BS	Tetrachloro-m-xylene	1	20	20.4	102		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	21.7	108		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.4	97		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	20.8	104		30 (32)	150 (175)
I.BLK-PP067758.D	PIBLK-PP067758.D	Tetrachloro-m-xylene	1	20	22.0	110		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	23.9	119		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	21.1	105		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	23.1	116		70 (60)	130 (140)
P4385-02	SP-1	Tetrachloro-m-xylene	1	20	14.2	71		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	14.2	71		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	14.0	70		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	13.8	69		30 (32)	150 (175)
P4385-04	SP-2	Tetrachloro-m-xylene	1	20	20.9	105		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	20.9	104		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	20.6	103		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	20.3	102		30 (32)	150 (175)
P4385-06	SP-3	Tetrachloro-m-xylene	1	20	20.3	101		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	20.8	104		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.6	98		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	20.0	100		30 (32)	150 (175)
P4385-08	SP-4	Tetrachloro-m-xylene	1	20	21.9	109		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	23.1	115		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	21.2	106		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	21.8	109		30 (32)	150 (175)
P4385-10	SP-5	Tetrachloro-m-xylene	1	20	20.6	103		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	21.3	106		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.8	99		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	20.3	101		30 (32)	150 (175)
P4385-12	SP-6	Tetrachloro-m-xylene	1	20	21.4	107		30 (32)	150 (144)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: **P4385**

Client: **Scheideler Excavating Co. Inc.**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
P4385-12	SP-6	Decachlorobiphenyl	1	20	21.4	107		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	21.4	107		30 (32)	150 (144)
P4385-14	SP-7	Decachlorobiphenyl	2	20	20.4	102		30 (32)	150 (175)
		Tetrachloro-m-xylene	1	20	14.7	74		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	15.3	77		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	14.3	71		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	14.9	75		30 (32)	150 (175)
P4385-16	SP-8	Tetrachloro-m-xylene	1	20	19.3	97		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	20.0	100		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.5	98		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	18.8	94		30 (32)	150 (175)
P4385-18	SP-9	Tetrachloro-m-xylene	1	20	20.4	102		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	20.8	104		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	20.0	100		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	19.3	96		30 (32)	150 (175)
P4385-20	SP-10	Tetrachloro-m-xylene	1	20	20.9	104		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	21.1	105		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	20.4	102		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	19.6	98		30 (32)	150 (175)
P4385-20MS	SP-10MS	Tetrachloro-m-xylene	1	20	21.6	108		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	20.7	104		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.5	97		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	19.0	95		30 (32)	150 (175)
P4385-20MSD	SP-10MSD	Tetrachloro-m-xylene	1	20	21.7	108		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	20.7	103		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.1	96		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	18.9	95		30 (32)	150 (175)
I.BLK-PP067772.D	PIBLK-PP067772.D	Tetrachloro-m-xylene	1	20	22.6	113		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	23.6	118		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	21.6	108		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	22.5	113		70 (60)	130 (140)