

Surrogate Summary

SDG No.: **P4385**

Client: **Scheideler Excavating Co. Inc.**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL091953.D	PIBLK-PL091953.D	Decachlorobiphenyl	1	20	19.6	98		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	19.0	95		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	18.7	94		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	18.3	92		30 (77)	150 (126)
I.BLK-PL092329.D	PIBLK-PL092329.D	Decachlorobiphenyl	1	20	23.8	119		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	21.3	107		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	22.1	110		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	20.6	103		30 (77)	150 (126)
PB164067BL	PB164067BL	Decachlorobiphenyl	1	20	19.6	98		30 (10)	150 (148)
		Tetrachloro-m-xylene	1	20	19.0	95		30 (10)	150 (159)
		Decachlorobiphenyl	2	20	19.6	98		30 (10)	150 (148)
		Tetrachloro-m-xylene	2	20	18.1	90		30 (10)	150 (159)
PB164067BS	PB164067BS	Decachlorobiphenyl	1	20	22.0	110		30 (10)	150 (148)
		Tetrachloro-m-xylene	1	20	20.7	103		30 (10)	150 (159)
		Decachlorobiphenyl	2	20	21.3	106		30 (10)	150 (148)
		Tetrachloro-m-xylene	2	20	19.9	100		30 (10)	150 (159)
P4385-02	SP-1	Decachlorobiphenyl	1	20	12.6	63		30 (10)	150 (148)
		Tetrachloro-m-xylene	1	20	13.2	66		30 (10)	150 (159)
		Decachlorobiphenyl	2	20	11.8	59		30 (10)	150 (148)
		Tetrachloro-m-xylene	2	20	12.9	64		30 (10)	150 (159)
P4385-04	SP-2	Decachlorobiphenyl	1	20	19.7	98		30 (10)	150 (148)
		Tetrachloro-m-xylene	1	20	20.0	100		30 (10)	150 (159)
		Decachlorobiphenyl	2	20	18.0	90		30 (10)	150 (148)
		Tetrachloro-m-xylene	2	20	20.1	101		30 (10)	150 (159)
P4385-04MS	SP-2MS	Decachlorobiphenyl	1	20	16.9	85		30 (10)	150 (148)
		Tetrachloro-m-xylene	1	20	16.4	82		30 (10)	150 (159)
		Decachlorobiphenyl	2	20	15.7	78		30 (10)	150 (148)
		Tetrachloro-m-xylene	2	20	16.3	81		30 (10)	150 (159)
P4385-04MSD	SP-2MSD	Decachlorobiphenyl	1	20	16.8	84		30 (10)	150 (148)
		Tetrachloro-m-xylene	1	20	16.6	83		30 (10)	150 (159)
		Decachlorobiphenyl	2	20	15.6	78		30 (10)	150 (148)
		Tetrachloro-m-xylene	2	20	15.9	79		30 (10)	150 (159)
P4385-08	SP-4	Decachlorobiphenyl	1	20	21.3	106		30 (10)	150 (148)
		Tetrachloro-m-xylene	1	20	20.9	104		30 (10)	150 (159)
		Decachlorobiphenyl	2	20	20.0	100		30 (10)	150 (148)
		Tetrachloro-m-xylene	2	20	20.4	102		30 (10)	150 (159)
P4385-10	SP-5	Decachlorobiphenyl	1	20	13.3	67		30 (10)	150 (148)
		Tetrachloro-m-xylene	1	20	12.6	63		30 (10)	150 (159)
		Decachlorobiphenyl	2	20	12.8	64		30 (10)	150 (148)
		Tetrachloro-m-xylene	2	20	12.1	60		30 (10)	150 (159)
P4385-14	SP-7	Decachlorobiphenyl	1	20	14.2	71		30 (10)	150 (148)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: **P4385**

Client: **Scheideler Excavating Co. Inc.**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
P4385-14	SP-7	Tetrachloro-m-xylene	1	20	14.0	70		30 (10)	150 (159)
		Decachlorobiphenyl	2	20	13.3	66		30 (10)	150 (148)
P4385-16	SP-8	Tetrachloro-m-xylene	2	20	13.1	65		30 (10)	150 (159)
		Decachlorobiphenyl	1	20	13.7	69		30 (10)	150 (148)
		Tetrachloro-m-xylene	1	20	13.3	66		30 (10)	150 (159)
		Decachlorobiphenyl	2	20	12.2	61		30 (10)	150 (148)
		Tetrachloro-m-xylene	2	20	12.5	62		30 (10)	150 (159)
P4385-20	SP-10	Decachlorobiphenyl	1	20	14.2	71		30 (10)	150 (148)
		Tetrachloro-m-xylene	1	20	13.4	67		30 (10)	150 (159)
		Decachlorobiphenyl	2	20	12.8	64		30 (10)	150 (148)
		Tetrachloro-m-xylene	2	20	13.1	66		30 (10)	150 (159)
P4385-06	SP-3	Decachlorobiphenyl	1	20	19.1	95		30 (10)	150 (148)
		Tetrachloro-m-xylene	1	20	20.0	100		30 (10)	150 (159)
		Decachlorobiphenyl	2	20	17.3	86		30 (10)	150 (148)
		Tetrachloro-m-xylene	2	20	19.5	97		30 (10)	150 (159)
P4385-18	SP-9	Decachlorobiphenyl	1	20	19.5	97		30 (10)	150 (148)
		Tetrachloro-m-xylene	1	20	20.0	100		30 (10)	150 (159)
		Decachlorobiphenyl	2	20	17.3	86		30 (10)	150 (148)
		Tetrachloro-m-xylene	2	20	19.7	99		30 (10)	150 (159)
P4385-12	SP-6	Decachlorobiphenyl	1	20	18.7	93		30 (10)	150 (148)
		Tetrachloro-m-xylene	1	20	20.0	100		30 (10)	150 (159)
		Decachlorobiphenyl	2	20	16.4	82		30 (10)	150 (148)
		Tetrachloro-m-xylene	2	20	19.5	98		30 (10)	150 (159)
I.BLK-PL092346.D	PIBLK-PL092346.D	Decachlorobiphenyl	1	20	21.1	105		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	21.4	107		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	19.7	99		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	21.2	106		30 (77)	150 (126)