

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4385

Client: Scheideler Excavating Co. Inc.

Analytical Method: 8081B

DataFile : PL092336.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	RPD		Limits		
			Result	Result			Qual	RPD	Qual	Low	High
Client Sample ID: P4385-04MS	SP-2MS										
	alpha-BHC	17.5	0	15.6	ug/kg	89			30 (60)	150 (144)	
	beta-BHC	17.5	0	16.0	ug/kg	91			30 (54)	150 (143)	
	delta-BHC	17.5	0	14.7	ug/kg	84			30 (47)	150 (144)	
	gamma-BHC (Lindane)	17.5	0	15.2	ug/kg	87			30 (61)	150 (140)	
	Heptachlor	17.5	0	16.0	ug/kg	91			30 (63)	150 (135)	
	Aldrin	17.5	0	15.3	ug/kg	87			30 (49)	150 (139)	
	Heptachlor epoxide	17.5	0	15.5	ug/kg	89			30 (32)	150 (180)	
	Endosulfan I	17.5	0	15.6	ug/kg	89			30 (56)	150 (142)	
	Dieldrin	17.5	0.94	16.7	ug/kg	90			30 (47)	150 (161)	
	4,4'-DDE	17.5	30.6	43.0	ug/kg	71			30 (55)	150 (136)	
	Endrin	17.5	0	16.9	ug/kg	97			30 (57)	150 (139)	
	Endosulfan II	17.5	0	16.0	ug/kg	91			30 (40)	150 (163)	
	4,4'-DDD	17.5	4	19.2	ug/kg	87			30 (37)	150 (192)	
	Endosulfan sulfate	17.5	0	16.1	ug/kg	92			30 (62)	150 (139)	
	4,4'-DDT	17.5	26	39.9	ug/kg	79			30 (51)	150 (146)	
	Methoxychlor	17.5	0	16.3	ug/kg	93			30 (54)	150 (136)	
	Endrin ketone	17.5	0	21.2	ug/kg	121			30 (60)	150 (129)	
	Endrin aldehyde	17.5	0	16.3	ug/kg	93			30 (59)	150 (132)	
	alpha-Chlordane	17.5	0	15.9	ug/kg	91			30 (30)	150 (192)	
gamma-Chlordane	17.5	0	16.4	ug/kg	94			30 (44)	150 (175)		