

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: SCHE03

Lab Code: CHEM Case No.: P4385 SAS No.: P4385 SDG No.: P4385

Instrument ID: BNA\_F Calibration Date/Time: 10/22/2024 14:28

Lab File ID: BF139927.D Init. Calib. Date(s): 10/18/2024 10/18/2024

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:27 13:46

GC Column: DB-UI ID: 0.18 (mm)

| COMPOUND                    | RRF   | RRF040 | MIN<br>RRF | %D   | MAX%D |
|-----------------------------|-------|--------|------------|------|-------|
| 2-Fluorophenol              | 1.278 | 1.238  |            | -3.1 |       |
| Benzaldehyde                | 0.931 | 0.957  |            | 2.8  |       |
| Phenol-d6                   | 1.655 | 1.570  |            | -5.1 |       |
| Phenol                      | 1.706 | 1.639  |            | -3.9 | 20.0  |
| bis(2-Chloroethyl)ether     | 1.319 | 1.272  |            | -3.6 |       |
| 2-Chlorophenol              | 1.306 | 1.287  |            | -1.5 |       |
| 2-Methylphenol              | 1.114 | 1.070  |            | -4.0 |       |
| 2,2-oxybis(1-Chloropropane) | 2.248 | 2.090  |            | -7.0 |       |
| Acetophenone                | 0.490 | 0.474  |            | -3.3 |       |
| 3+4-Methylphenols           | 1.424 | 1.367  |            | -4.0 |       |
| n-Nitroso-di-n-propylamine  | 1.004 | 0.934  | 0.050      | -7.0 |       |
| Nitrobenzene-d5             | 0.361 | 0.377  |            | 4.4  |       |
| Hexachloroethane            | 0.534 | 0.531  |            | -0.6 |       |
| Nitrobenzene                | 0.396 | 0.401  |            | 1.3  |       |
| Isophorone                  | 0.685 | 0.658  |            | -3.9 |       |
| 2-Nitrophenol               | 0.149 | 0.178  |            | 19.5 | 20.0  |
| 2,4-Dimethylphenol          | 0.249 | 0.243  |            | -2.4 |       |
| bis(2-Chloroethoxy)methane  | 0.415 | 0.406  |            | -2.2 |       |
| 2,4-Dichlorophenol          | 0.283 | 0.285  |            | 0.7  | 20.0  |
| Naphthalene                 | 1.031 | 1.017  |            | -1.4 |       |
| 4-Chloroaniline             | 0.352 | 0.346  |            | -1.7 |       |
| Hexachlorobutadiene         | 0.197 | 0.204  |            | 3.6  | 20.0  |
| Caprolactam                 | 0.090 | 0.086  |            | -4.4 |       |
| 4-Chloro-3-methylphenol     | 0.314 | 0.306  |            | -2.5 | 20.0  |
| 2-Methylnaphthalene         | 0.632 | 0.625  |            | -1.1 |       |
| Hexachlorocyclopentadiene   | 0.188 | 0.217  | 0.050      | 15.4 |       |
| 2,4,6-Trichlorophenol       | 0.382 | 0.396  |            | 3.7  | 20.0  |
| 2-Fluorobiphenyl            | 1.210 | 1.202  |            | -0.7 |       |
| 2,4,5-Trichlorophenol       | 0.394 | 0.423  |            | 7.4  |       |
| 1,1-Biphenyl                | 1.392 | 1.411  |            | 1.4  |       |
| 2-Chloronaphthalene         | 1.121 | 1.140  |            | 1.7  |       |
| 2-Nitroaniline              | 0.343 | 0.368  |            | 7.3  |       |
| Dimethylphthalate           | 1.246 | 1.230  |            | -1.3 |       |
| Acenaphthylene              | 1.621 | 1.627  |            | 0.4  |       |
| 2,6-Dinitrotoluene          | 0.270 | 0.289  |            | 7.0  |       |
| 3-Nitroaniline              | 0.278 | 0.287  |            | 3.2  |       |
| Acenaphthene                | 1.049 | 1.066  |            | 1.6  | 20.0  |
| 2,4-Dinitrophenol           | 0.093 | 0.132  | 0.050      | 41.9 |       |
| 4-Nitrophenol               | 0.214 | 0.223  | 0.050      | 4.2  |       |

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: SCHE03

Lab Code: CHEM Case No.: P4385 SAS No.: P4385 SDG No.: P4385

Instrument ID: BNA\_F Calibration Date/Time: 10/22/2024 14:28

Lab File ID: BF139927.D Init. Calib. Date(s): 10/18/2024 10/18/2024

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:27 13:46

GC Column: DB-UI ID: 0.18 (mm)

| COMPOUND                   | RRF   | RRF040 | MIN<br>RRF | %D    | MAX%D |
|----------------------------|-------|--------|------------|-------|-------|
| Dibenzofuran               | 1.505 | 1.499  |            | -0.4  |       |
| 2,4-Dinitrotoluene         | 0.332 | 0.373  |            | 12.3  |       |
| Diethylphthalate           | 1.223 | 1.185  |            | -3.1  |       |
| 4-Chlorophenyl-phenylether | 0.579 | 0.580  |            | 0.2   |       |
| Fluorene                   | 1.146 | 1.117  |            | -2.5  |       |
| 4-Nitroaniline             | 0.263 | 0.268  |            | 1.9   |       |
| 4,6-Dinitro-2-methylphenol | 0.082 | 0.112  |            | 36.6  |       |
| n-Nitrosodiphenylamine     | 0.599 | 0.598  |            | -0.2  | 20.0  |
| 2,4,6-Tribromophenol       | 0.187 | 0.201  |            | 7.5   |       |
| 4-Bromophenyl-phenylether  | 0.206 | 0.216  |            | 4.9   |       |
| Hexachlorobenzene          | 0.232 | 0.239  |            | 3.0   |       |
| Atrazine                   | 0.162 | 0.155  |            | -4.3  |       |
| Pentachlorophenol          | 0.141 | 0.160  |            | 13.5  | 20.0  |
| Phenanthrene               | 0.945 | 0.925  |            | -2.1  |       |
| Anthracene                 | 0.922 | 0.916  |            | -0.7  |       |
| Carbazole                  | 0.857 | 0.810  |            | -5.5  |       |
| Di-n-butylphthalate        | 0.990 | 0.919  |            | -7.2  |       |
| Fluoranthene               | 0.955 | 0.875  |            | -8.5  | 20.0  |
| Pyrene                     | 1.751 | 1.852  |            | 5.8   |       |
| Terphenyl-d14              | 1.227 | 1.269  |            | 3.4   |       |
| Butylbenzylphthalate       | 0.525 | 0.539  |            | 2.7   |       |
| 3,3-Dichlorobenzidine      | 0.380 | 0.452  |            | 18.9  |       |
| Benzo (a) anthracene       | 1.302 | 1.303  |            | 0.1   |       |
| Chrysene                   | 1.194 | 1.167  |            | -2.3  |       |
| Bis(2-ethylhexyl)phthalate | 0.589 | 0.693  |            | 17.7  |       |
| Di-n-octyl phthalate       | 1.071 | 1.312  |            | 22.5  | 20.0  |
| Benzo (b) fluoranthene     | 1.218 | 1.229  |            | 0.9   |       |
| Benzo (k) fluoranthene     | 1.051 | 0.899  |            | -14.5 |       |
| Benzo (a) pyrene           | 1.002 | 1.008  |            | 0.6   | 20.0  |
| Indeno (1,2,3-cd) pyrene   | 1.287 | 1.437  |            | 11.7  |       |
| Dibenzo (a,h) anthracene   | 1.073 | 1.186  |            | 10.5  |       |
| Benzo (g,h,i) perylene     | 1.072 | 1.213  |            | 13.2  |       |
| 1,2,4,5-Tetrachlorobenzene | 0.540 | 0.557  |            | 3.1   |       |
| 1,4-Dioxane                | 0.596 | 0.561  |            | -5.9  | 20.0  |
| 2,3,4,6-Tetrachlorophenol  | 0.309 | 0.321  |            | 3.9   |       |

All other compounds must meet a minimum RRF of 0.010.