


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102124\
Data File : BF139891.D
Acq On : 21 Oct 2024 09:28
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Oct 18 15:07:50 2024






AutoFind: Scans 1627, 1628, 1629; Background Corrected with Scan 1621

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.9	119787	PASS
68	69	0.00	2	1.7	1765	PASS
69	198	0.00	100	38.2	104142	PASS
70	69	0.00	2	0.6	599	PASS
127	198	10	80	47.1	128648	PASS
197	198	0.00	2	0.4	1218	PASS
198	198	100	100	100.0	272917	PASS
199	198	5	9	6.6	18009	PASS
275	198	10	60	27.7	75493	PASS
365	198	1	100	3.5	9452	PASS
441	198	0.01	100	15.4	41963	PASS
442	442	50	100	100.0	268416	PASS
443	442	15	24	19.4	51981	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102124\
Data File : BF139891.D
Acq On : 21 Oct 2024 09:28
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Oct 21 11:06:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration



TIC: BF139891.D\data.ms

(70) Pentachlorophenol (C)

11.216min (+ 0.006) 62101.95 ng




response 401749

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.00	63.39
264.00	63.20	62.51
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102124\
Data File : BF139891.D
Acq On : 21 Oct 2024 09:28
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Oct 21 11:06:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration



TIC: BF139891.D\data.ms

(77) Benzidine

12.757min (+ 0.012) 76266.80 ng

response 1925021

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.00	14.03
183.00	11.80	11.90
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/21/2024	BNA_F	<u>BF139891.D</u>
Compound Name	Response	Retention Time
DDT	854790	13.58
DDD	11947	13.31
DDE	336	12.939
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
12283	867073	1.42

Instrument :
BNA_F
ClientSampleId :
DFTPP