

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: SCHE03
 Lab Code: CHEM Case No.: P4385 SAS No.: P4385 SDG NO.: P4385
 Lab File ID: VY019882.D Date Analyzed: 10/14/2024
 Instrument ID: MSVOA_Y Time Analyzed: 09:52
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	286716	7.71	498763	8.62	439404	11.41
UPPER LIMIT	573432	8.207	997526	9.116	878808	11.914
LOWER LIMIT	143358	7.207	249382	8.116	219702	10.914
EPA SAMPLE NO.						
SP-6	255373	7.71	522145	8.62	475654	11.41
SP-7	246979	7.71	513227	8.62	465047	11.41
VY1014SBL01	272092	7.71	541333	8.62	470368	11.41
VY1014SBS01	286988	7.71	511012	8.62	442973	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.