

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: SCHE03
Lab Code: CHEM Case No.: P4385 SAS No.: P4385 SDG NO.: P4385
Lab File ID: VY019862.D Date Analyzed: 10/11/2024
Instrument ID: MSVOA_Y Time Analyzed: 10:11
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	247612	13.346				
UPPER LIMIT	495224	13.846				
LOWER LIMIT	123806	12.846				
EPA SAMPLE NO.						
SP-1	182418	13.35				
SP-2	173582	13.35				
SP-3	148777	13.35				
SP-4	157772	13.35				
SP-5	175004	13.35				
SP-8	154054	13.35				
SP-9	181975	13.35				
SP-10	170163	13.35				
VY1011SBL01	181683	13.35				
VY1011SBS01	243672	13.35				
VY1011SBSD01	227990	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.