

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: SCHE03
 Lab Code: CHEM Case No.: P4385 SAS No.: P4385 SDG NO.: P4385
 Lab File ID: VY019882.D Date Analyzed: 10/14/2024
 Instrument ID: MSVOA_Y Time Analyzed: 09:52
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	212252	13.347				
UPPER LIMIT	424504	13.847				
LOWER LIMIT	106126	12.847				
EPA SAMPLE NO.						
SP-6	169983	13.35				
SP-7	168651	13.35				
VY1014SBL01	156619	13.35				
VY1014SBS01	215154	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.