


LB133065

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.019	0.019	-21	0.000	-21	0				22
0.2	0.200	0.243	0.043	916	0.0 %	916					5
2.5	2.500	2.383	-0.117	9889	0.0 %	9889					2
5.0	5.000	5.085	0.085	21216	0.0 %	21216					1
7.5	7.500	7.447	-0.053	31115	0.0 %	31115					0
10.0	10.000	10.023	0.023	41914	0.0 %	41914					

LB133065 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Oct 2024 08:54:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-21	-	0.0000	7470A	PPB	23 Oct 2024 09:26:45	S	Std
0.2 - 1	50.2	916	-	0.2000	7470A	PPB	23 Oct 2024 09:29:02	S	Std
2.5 - 1	52.5	9889	-	2.5000	7470A	PPB	23 Oct 2024 09:31:19	S	Std
5.0 - 1	55	21216	-	5.0000	7470A	PPB	23 Oct 2024 09:33:36	S	Std
7.5 - 1	57.5	31115	-	7.5000	7470A	PPB	23 Oct 2024 09:35:51	S	Std
10.0 - 1	510	41914	-	10.0000	7470A	PPB	23 Oct 2024 09:40:51	S	Std
ICV48 - 1	ICV48	16495	3.9591	-	7470A	PPB	23 Oct 2024 09:44:25	U	SMPL
ICB48 - 1	ICB48	-363	-0.0625	-	7470A	PPB	23 Oct 2024 09:46:40	U	SMPL
CCV56 - 1	CCV56	20937	5.0188	-	7470A	PPB	23 Oct 2024 09:48:56	U	SMPL
CCB56 - 1	CCB56	-317	-0.0515	-	7470A	PPB	23 Oct 2024 09:51:11	U	SMPL
CRA - 1	CRA	810	0.2173	-	7470A	PPB	23 Oct 2024 09:53:27	U	SMPL
HighStd - 1	HighStd	42281	10.1105	-	7470A	PPB	23 Oct 2024 09:55:42	U	SMPL
ChkStd - 1	ChkStd	29304	7.0148	-	7470A	PPB	23 Oct 2024 10:03:45	U	SMPL
PB164306BL - 1	PB164306BL	-365	-0.0630	-	7470A	PPB	23 Oct 2024 10:06:01	U	SMPL
PB164306BS - 1	PB164306BS	14663	3.5221	-	7470A	PPB	23 Oct 2024 10:08:18	U	SMPL
P4397-06 - 1	WB-301-BOT	-211	-0.0262	-	7470A	PPB	23 Oct 2024 10:10:34	U	SMPL
P4443-05 - 1	OG-315-HR-502-COMP-29	-117	-0.0038	-	7470A	PPB	23 Oct 2024 10:12:52	U	SMPL
P4443-10 - 1	OG-315-HR-502-COMP-30	-3	0.0234	-	7470A	PPB	23 Oct 2024 10:15:08	U	SMPL
P4458-02 - 1	280517	-43	0.0139	-	7470A	PPB	23 Oct 2024 10:17:28	U	SMPL
P4460-04 - 1	WB-303-BOT	-90	0.0026	-	7470A	PPB	23 Oct 2024 10:19:46	U	SMPL
CCV57 - 1	CCV57	22150	5.3081	-	7470A	PPB	23 Oct 2024 10:22:01	U	SMPL
CCB57 - 1	CCB57	-386	-0.0680	-	7470A	PPB	23 Oct 2024 10:24:16	U	SMPL
P4460-04DUP - 1	WB-303-BOTDUP	-94	0.0017	-	7470A	PPB	23 Oct 2024 10:26:33	U	SMPL
P4460-04MS - 1	WB-303-BOTMS	13457	3.2344	-	7470A	PPB	23 Oct 2024 10:28:48	U	SMPL
P4460-04MSD - 1	WB-303-BOTMSD	16951	3.8293	-	7470A	PPB	23 Oct 2024 10:31:03	U	SMPL
PB164321BL - 1	PB164321BL	-344	-0.0579	-	7470A	PPB	23 Oct 2024 10:33:20	U	SMPL
PB164321BS - 1	PB164321BS	16147	3.8761	-	7470A	PPB	23 Oct 2024 10:35:38	U	SMPL
P4467-04 - 1	TP-1	-260	-0.0379	-	7470A	PPB	23 Oct 2024 10:37:54	U	SMPL
P4468-02 - 1	ETGI-331	-83	0.0043	-	7470A	PPB	23 Oct 2024 10:40:12	U	SMPL
P4468-04 - 1	ETGI-329	-75	0.0062	-	7470A	PPB	23 Oct 2024 10:42:29	U	SMPL
P4468-06 - 1	ETGI-345	2	0.0246	-	7470A	PPB	23 Oct 2024 10:44:45	U	SMPL
P4472-04 - 1	BP-F-28	-27	0.0177	-	7470A	PPB	23 Oct 2024 10:47:01	U	SMPL
CCV58 - 1	CCV58	22679	5.4343	-	7470A	PPB	23 Oct 2024 10:49:16	U	SMPL
CCB58 - 1	CCB58	-316	-0.0513	-	7470A	PPB	23 Oct 2024 10:51:32	U	SMPL
P4472-08 - 1	BP-F-6	-30	0.0170	-	7470A	PPB	23 Oct 2024 10:53:49	U	SMPL
P4472-08DUP - 1	BP-F-6DUP	-6	0.0227	-	7470A	PPB	23 Oct 2024 10:56:04	U	SMPL
P4472-08MS - 1	BP-F-6MS	14646	3.5180	-	7470A	PPB	23 Oct 2024 10:58:20	U	SMPL
P4472-08MSD - 1	BP-F-6MSD	14059	3.3780	-	7470A	PPB	23 Oct 2024 11:00:35	U	SMPL
PB164261TB - 1	PB164261TB	-352	-0.0599	-	7470A	PPB	23 Oct 2024 11:02:53	U	SMPL
PB164301TB - 1	PB164301TB	-66	0.0084	-	7470A	PPB	23 Oct 2024 11:05:15	U	SMPL
P4460-04LX5 - 1		31	0.0315	-	7470A	PPB	23 Oct 2024 11:07:32	U	SMPL
P4460-04A - 1		16402	3.9369	-	7470A	PPB	23 Oct 2024 11:09:48	U	SMPL
P4472-08LX5 - 1		-256	-0.0370	-	7470A	PPB	23 Oct 2024 11:12:05	U	SMPL
P4472-08A - 1		17250	4.1392	-	7470A	PPB	23 Oct 2024 11:14:22	U	SMPL
CCV59 - 1	CCV59	22140	5.3057	-	7470A	PPB	23 Oct 2024 11:16:37	U	SMPL
CCB59 - 1	CCB59	-458	-0.0851	-	7470A	PPB	23 Oct 2024 11:18:54	U	SMPL