

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB133065

Review By	jaswal	Review On	10/23/2024 11:18:30 PM
Supervise By	mohan	Supervise On	10/23/2024 11:33:40 PM

STD. NAME	STD REF.#
ICAL Standard	MP82866,MP82868,MP82869,MP82870,MP82871,MP82872
ICV Standard	MP82873
CCV Standard	MP82875
ICSA Standard	
CRI Standard	MP82877
LCS Standard	
Chk Standard	MP82874,MP82876,MP82883,MP82885

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/23/24 09:26		Mohan	OK
2	S0.2	S0.2	CAL2	10/23/24 09:29		Mohan	OK
3	S2.5	S2.5	CAL3	10/23/24 09:31		Mohan	OK
4	S5	S5	CAL4	10/23/24 09:33		Mohan	OK
5	S7.5	S7.5	CAL5	10/23/24 09:35		Mohan	OK
6	S10	S10	CAL6	10/23/24 09:40		Mohan	OK
7	ICV48	ICV48	ICV	10/23/24 09:44		Mohan	OK
8	ICB48	ICB48	ICB	10/23/24 09:46		Mohan	OK
9	CCV56	CCV56	CCV	10/23/24 09:48		Mohan	OK
10	CCB56	CCB56	CCB	10/23/24 09:51		Mohan	OK
11	CRA	CRA	CRDL	10/23/24 09:53		Mohan	OK
12	HighStd	HighStd	HIGH STD	10/23/24 09:55		Mohan	OK
13	ChkStd	ChkStd	SAM	10/23/24 10:03		Mohan	OK
14	PB164306BL	PB164306BL	MB	10/23/24 10:06		Mohan	OK
15	PB164306BS	PB164306BS	LCS	10/23/24 10:08		Mohan	OK
16	P4397-06	WB-301-BOT	SAM	10/23/24 10:10		Mohan	OK
17	P4443-05	OG-315-HR-502-COM	SAM	10/23/24 10:12		Mohan	OK
18	P4443-10	OG-315-HR-502-COM	SAM	10/23/24 10:15		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB133065

Review By	jaswal	Review On	10/23/2024 11:18:30 PM
Supervise By	mohan	Supervise On	10/23/2024 11:33:40 PM

STD. NAME	STD REF.#
ICAL Standard	MP82866,MP82868,MP82869,MP82870,MP82871,MP82872
ICV Standard	MP82873
CCV Standard	MP82875
ICSA Standard	
CRI Standard	MP82877
LCS Standard	
Chk Standard	MP82874,MP82876,MP82883,MP82885

19	P4458-02	280517	SAM	10/23/24 10:17		Mohan	OK
20	P4460-04	WB-303-BOT	SAM	10/23/24 10:19		Mohan	OK
21	CCV57	CCV57	CCV	10/23/24 10:22		Mohan	OK
22	CCB57	CCB57	CCB	10/23/24 10:24		Mohan	OK
23	P4460-04DUP	WB-303-BOTDUP	DUP	10/23/24 10:26		Mohan	OK
24	P4460-04MS	WB-303-BOTMS	MS	10/23/24 10:28		Mohan	OK
25	P4460-04MSD	WB-303-BOTMSD	MSD	10/23/24 10:31		Mohan	OK
26	PB164321BL	PB164321BL	MB	10/23/24 10:33		Mohan	OK
27	PB164321BS	PB164321BS	LCS	10/23/24 10:35		Mohan	OK
28	P4467-04	TP-1	SAM	10/23/24 10:37		Mohan	OK
29	P4468-02	ETGI-331	SAM	10/23/24 10:40		Mohan	OK
30	P4468-04	ETGI-329	SAM	10/23/24 10:42		Mohan	OK
31	P4468-06	ETGI-345	SAM	10/23/24 10:44		Mohan	OK
32	P4472-04	BP-F-28	SAM	10/23/24 10:47		Mohan	OK
33	CCV58	CCV58	CCV	10/23/24 10:49		Mohan	OK
34	CCB58	CCB58	CCB	10/23/24 10:51		Mohan	OK
35	P4472-08	BP-F-6	SAM	10/23/24 10:53		Mohan	OK
36	P4472-08DUP	BP-F-6DUP	DUP	10/23/24 10:56		Mohan	OK
37	P4472-08MS	BP-F-6MS	MS	10/23/24 10:58		Mohan	OK
38	P4472-08MSD	BP-F-6MSD	MSD	10/23/24 11:00		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB133065

Review By	jaswal	Review On	10/23/2024 11:18:30 PM
Supervise By	mohan	Supervise On	10/23/2024 11:33:40 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82866,MP82868,MP82869,MP82870,MP82871,MP82872		
ICV Standard	MP82873		
CCV Standard	MP82875		
ICSA Standard			
CRI Standard	MP82877		
LCS Standard			
Chk Standard	MP82874,MP82876,MP82883,MP82885		

39	PB164261TB	PB164261TB	MB	10/23/24 11:02		Mohan	OK
40	PB164301TB	PB164301TB	MB	10/23/24 11:05		Mohan	OK
41	P4460-04L	WB-303-BOTL	SD	10/23/24 11:07		Mohan	OK
42	P4460-04A	WB-303-BOTA	PS	10/23/24 11:09		Mohan	OK
43	P4472-08L	BP-F-6L	SD	10/23/24 11:12		Mohan	OK
44	P4472-08A	BP-F-6A	PS	10/23/24 11:14		Mohan	OK
45	CCV59	CCV59	CCV	10/23/24 11:16		Mohan	OK
46	CCB59	CCB59	CCB	10/23/24 11:18		Mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133086

Review By	kareem	Review On	10/25/2024 3:10:26 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:07:20 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712
ICV Standard	mp82485
CCV Standard	MP82488
ICSA Standard	MP82486 MP82487
CRI Standard	MP82712
LCS Standard	
Chk Standard	MP82491 MP82492

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/23/24 14:24		Kareem	OK
2	S1	S1	CAL2	10/23/24 14:28		Kareem	OK
3	S2	S2	CAL3	10/23/24 14:32		Kareem	OK
4	S3	S3	CAL4	10/23/24 14:37		Kareem	OK
5	S4	S4	CAL5	10/23/24 14:41		Kareem	OK
6	S5	S5	CAL6	10/23/24 14:45		Kareem	OK
7	ICV01	ICV01	ICV	10/23/24 14:59		Kareem	OK
8	LLICV01	LLICV01	LLICV	10/23/24 15:11		Kareem	OK
9	ICB01	ICB01	ICB	10/23/24 15:16		Kareem	OK
10	CRI01	CRI01	CRDL	10/23/24 15:22		Kareem	OK
11	ICSA01	ICSA01	ICSA	10/23/24 15:26		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	10/23/24 15:32		Kareem	OK
13	ICSADL	ICSADL	ICSA	10/23/24 15:36		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	10/23/24 15:40		Kareem	OK
15	CCV01	CCV01	CCV	10/23/24 15:57		Kareem	OK
16	CCB01	CCB01	CCB	10/23/24 16:02		Kareem	OK
17	P4458-02	280517	SAM	10/23/24 16:06		Kareem	OK
18	P4458-01	280517	SAM	10/23/24 16:10		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133086

Review By	kareem	Review On	10/25/2024 3:10:26 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:07:20 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712
ICV Standard	mp82485
CCV Standard	MP82488
ICSA Standard	MP82486 MP82487
CRI Standard	MP82712
LCS Standard	
Chk Standard	MP82491 MP82492

19	P4443-05	OG-315-HR-502-COM	SAM	10/23/24 16:15		Kareem	OK
20	P4443-10	OG-315-HR-502-COM	SAM	10/23/24 16:19		Kareem	OK
21	P4443-01	OG-315-HR-502-COM	SAM	10/23/24 16:23		Kareem	OK
22	P4443-06	OG-315-HR-502-COM	SAM	10/23/24 16:28		Kareem	OK
23	P4452-01	ETGI-285	SAM	10/23/24 16:32	Ca,Cr high	Kareem	Dilution
24	P4452-01DL	ETGI-285DL	SAM	10/23/24 16:44	5x for Ca,Cr	Kareem	Confirms
25	CCV02	CCV02	CCV	10/23/24 16:48		Kareem	OK
26	CCB02	CCB02	CCB	10/23/24 16:52		Kareem	OK
27	P4397-06	WB-301-BOT	SAM	10/23/24 16:59		Kareem	OK
28	P4460-04	WB-303-BOT	SAM	10/23/24 17:03		Kareem	OK
29	P4460-04DUP	WB-303-BOTDUP	DUP	10/23/24 17:08		Kareem	OK
30	P4460-04L	WB-303-BOTL	SD	10/23/24 17:12		Kareem	OK
31	P4460-04MS	WB-303-BOTMS	MS	10/23/24 17:17	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
32	P4460-04MSD	WB-303-BOTMSD	MSD	10/23/24 17:21	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
33	P4460-04A	WB-303-BOTA	PS	10/23/24 17:25	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
34	PB164261TB	PB164261TB	MB	10/23/24 17:29		Kareem	OK
35	PB164298BL	PB164298BL	MB	10/23/24 17:34		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133086

Review By	kareem	Review On	10/25/2024 3:10:26 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:07:20 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712
ICV Standard	mp82485
CCV Standard	MP82488
ICSA Standard	MP82486 MP82487
CRI Standard	MP82712
LCS Standard	
Chk Standard	MP82491 MP82492

36	PB164298BS	PB164298BS	LCS	10/23/24 17:38	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
37	CCV03	CCV03	CCV	10/23/24 17:42		Kareem	OK
38	CCB03	CCB03	CCB	10/23/24 17:46		Kareem	OK
39	P4456-01	PAD-10182024	SAM	10/23/24 17:51		Kareem	OK
40	P4460-02	WB-303-TOP	SAM	10/23/24 17:55		Kareem	OK
41	P4460-03	WB-303-BOT	SAM	10/23/24 17:59	MS-MSD Fail for more than 50% parameter	Kareem	OK
42	P4460-03DUP	WB-303-BOTDUP	DUP	10/23/24 18:04	MS-MSD Fail for more than 50% parameter	Kareem	OK
43	P4460-03L	WB-303-BOTL	SD	10/23/24 18:08	MS-MSD Fail for more than 50% parameter	Kareem	OK
44	P4460-03MS	WB-303-BOTMS	MS	10/23/24 18:12	MS-MSD Fail for more than 50% parameter	Kareem	OK
45	P4460-03MSD	WB-303-BOTMSD	MSD	10/23/24 18:16	MS-MSD Fail for more than 50% parameter	Kareem	OK
46	P4460-03A	WB-303-BOTA	PS	10/23/24 18:20	MS-MSD Fail for more than 50% parameter	Kareem	OK
47	PB164289BL	PB164289BL	MB	10/23/24 18:24	Fail for Al	Kareem	Not Ok
48	PB164289BS	PB164289BS	LCS	10/23/24 18:28	Fail for Al,Sb,As,Ba,Be,Cd,Pb,Mg, Mn,Se,Tl,V	Kareem	Not Ok
49	CCV04	CCV04	CCV	10/23/24 18:32		Kareem	OK
50	CCB04	CCB04	CCB	10/23/24 18:36		Kareem	OK
51	PB164244BL	PB164244BL	MB	10/23/24 18:41		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133086

Review By	kareem	Review On	10/25/2024 3:10:26 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:07:20 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712
ICV Standard	mp82485
CCV Standard	MP82488
ICSA Standard	MP82486 MP82487
CRI Standard	MP82712
LCS Standard	
Chk Standard	MP82491 MP82492

52	PB164244BS	PB164244BS	LCS	10/23/24 18:45	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
53	PB164195TB	PB164195TB	MB	10/23/24 18:49		Kareem	OK
54	PB164197TB	PB164197TB	MB	10/23/24 18:54		Kareem	OK
55	PB164248BL	PB164248BL	MB	10/23/24 18:58		Kareem	OK
56	PB164248BS	PB164248BS	LCS	10/23/24 19:03	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
57	PB164174BL	PB164174BL	MB	10/23/24 19:07		Kareem	OK
58	PB164174BS	PB164174BS	LCS	10/23/24 19:11	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
59	PB164221BL	PB164221BL	MB	10/23/24 19:15		Kareem	OK
60	PB164221BS	PB164221BS	LCS	10/23/24 19:19	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
61	CCV05	CCV05	CCV	10/23/24 19:23		Kareem	OK
62	CCB05	CCB05	CCB	10/23/24 19:28		Kareem	OK
63	PB164222BL	PB164222BL	MB	10/23/24 19:32		Kareem	OK
64	PB164222BS	PB164222BS	LCS	10/23/24 19:36	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
65	PB164220BL	PB164220BL	MB	10/23/24 19:40		Kareem	OK
66	PB164220BS	PB164220BS	LCS	10/23/24 19:45	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133086

Review By	kareem	Review On	10/25/2024 3:10:26 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:07:20 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712
ICV Standard	mp82485
CCV Standard	MP82488
ICSA Standard	MP82486 MP82487
CRI Standard	MP82712
LCS Standard	
Chk Standard	MP82491 MP82492

67	P4424-01	1-N-1	SAM	10/23/24 19:49		Kareem	OK
68	P4424-01DUP	1-N-1DUP	DUP	10/23/24 19:53		Kareem	OK
69	P4424-01L	1-N-1L	SD	10/23/24 19:58		Kareem	OK
70	P4424-01MS	1-N-1MS	MS	10/23/24 20:02	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
71	P4424-01MSD	1-N-1MSD	MSD	10/23/24 20:06	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
72	P4424-01A	1-N-1A	PS	10/23/24 20:11	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
73	CCV06	CCV06	CCV	10/23/24 20:15		Kareem	OK
74	CCB06	CCB06	CCB	10/23/24 20:21		Kareem	OK
75	P4424-02	1-S-1	SAM	10/23/24 20:25		Kareem	OK
76	P4424-03	1-B-1	SAM	10/23/24 20:30		Kareem	OK
77	P4397-04DL	WB-301-SWDL	SAM	10/23/24 20:47	Not Use	Kareem	Not Ok
78	P4397-04DUPDL	WB-301-SWDUPDL	DUP	10/23/24 20:51	Not Use	Kareem	Not Ok
79	P4397-04LDL	WB-301-SWLDL	SD	10/23/24 20:56	Not Use	Kareem	Not Ok
80	P4397-04MSDL	WB-301-SWMSDL	MS	10/23/24 21:00	Not Use	Kareem	Not Ok
81	P4397-04MSDDL	WB-301-SWMSDDL	MSD	10/23/24 21:05	Not Use	Kareem	Not Ok
82	LR1	LR1	HIGH STD	10/23/24 21:11		Kareem	OK
83	LR2	LR2	HIGH STD	10/23/24 21:16		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133086

Review By	kareem	Review On	10/25/2024 3:10:26 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:07:20 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712		
ICV Standard	mp82485		
CCV Standard	MP82488		
ICSA Standard	MP82486 MP82487		
CRI Standard	MP82712		
LCS Standard			
Chk Standard	MP82491 MP82492		

84	CCV07	CCV07	CCV	10/23/24 21:28		Kareem	OK
85	CCB07	CCB07	CCB	10/23/24 21:34		Kareem	OK
86	P4397-04ADL	WB-301-SWADL	PS	10/23/24 21:39	Not Use	Kareem	Not Ok
87	P4347-01	EFFLUENT-DAY-1-M	SAM	10/23/24 21:43		Kareem	OK
88	P4347-01DUP	EFFLUENT-DAY-1-M	DUP	10/23/24 21:47		Kareem	OK
89	P4347-01L	EFFLUENT-DAY-1-M	SD	10/23/24 21:52		Kareem	OK
90	P4347-01MS	EFFLUENT-DAY-1-M	MS	10/23/24 21:56	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
91	P4347-01MSD	EFFLUENT-DAY-1-M	MSD	10/23/24 22:04	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
92	P4347-01A	EFFLUENT-DAY-1-M	PS	10/23/24 22:08	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
93	PB164302BL	PB164302BL	MB	10/23/24 22:12		Kareem	OK
94	PB164302BS	PB164302BS	LCS	10/23/24 22:16	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
95	CCV08	CCV08	CCV	10/23/24 22:20		Kareem	OK
96	CCB08	CCB08	CCB	10/23/24 22:25		Kareem	OK