



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Portal Partners Tri-Venture **SDG No.:** P4397
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4397 **SAS No.:** P4397
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.22	0.2	109	40 - 160	CV	10/23/2024	09:53	LB133065
CRI01	Arsenic	19.7	20.0	99	40 - 160	P	10/23/2024	15:22	LB133086
	Barium	92.7	100	93	40 - 160	P	10/23/2024	15:22	LB133086
	Cadmium	5.38	6.0	90	40 - 160	P	10/23/2024	15:22	LB133086
	Chromium	10.0	10.0	100	40 - 160	P	10/23/2024	15:22	LB133086
	Lead	11.1	12.0	92	40 - 160	P	10/23/2024	15:22	LB133086
	Selenium	17.6	20.0	88	40 - 160	P	10/23/2024	15:22	LB133086
	Silver	10.1	10.0	101	40 - 160	P	10/23/2024	15:22	LB133086