

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Portal Partners Tri-Venture **SDG No.:** P4397
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4397 **SAS No.:** P4397
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	6.17			-20	20	10/23/2024	15:26	LB133086
	Barium	4.64	6.0	77	-94	106	10/23/2024	15:26	LB133086
	Cadmium	5.24	1.0	524	-5	7	10/23/2024	15:26	LB133086
	Chromium	58.7	52.0	113	42	62	10/23/2024	15:26	LB133086
	Lead	7.85			-12	12	10/23/2024	15:26	LB133086
	Selenium	-14.6			-20	20	10/23/2024	15:26	LB133086
	Silver	0.40			-10	10	10/23/2024	15:26	LB133086
ICSAB01	Arsenic	109	104	105	88.4	120	10/23/2024	15:32	LB133086
	Barium	477	537	89	437	637	10/23/2024	15:32	LB133086
	Cadmium	987	972	102	826	1120	10/23/2024	15:32	LB133086
	Chromium	574	542	106	460	624	10/23/2024	15:32	LB133086
	Lead	55.0	49.0	112	37	61	10/23/2024	15:32	LB133086
	Selenium	32.1	46.0	70	26	66	10/23/2024	15:32	LB133086
	Silver	206	201	102	170	232	10/23/2024	15:32	LB133086